



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

No. AIIMS/R/HS/BB/11/155/2021

Dated: - 08-03-2021

**Technical Corrigendum Notice**

**Tender ID. No. 2021\_IMSRP\_619933\_1**

**Dated: 02/03/2021**

With reference to the above Tender Id no., the following specifications of item that are needed to be revised which are as under:

S. No.	Title under which amendments made	Revised additional specifications
1	Technical Specification Of Red cell Filtration	<ul style="list-style-type: none"><li>i) Dockable Lab Side Red cell filter - Should provide 4 log reduction of leucocytes</li><li>ii) Should have attached 1PVC DEHP transfer bag (600 ml capacity).</li><li>iii) Pre filter 200micro meter</li><li>iv) Should be made of proven biocompatible material which is neutral.</li><li>v) Should Enable the Red cell concentrate filtration up to 24 hrs after collection at room temperature and up to 72 hours after collection and Storage at 4 °C</li><li>vi) Should be laboratory type filter.</li><li>vii) Two layered filter system for up to 90% Recovery of Red cell.</li><li>viii) By pass system to maintain the sterility and Recovery, no air vent.</li><li>ix) Should be sterilized by Beta radiation.</li><li>x) Performance as per AABB and European Guidelines.</li><li>xi) Three years' shelf life.</li><li>xii) Should meet all standards and have CE/BIS marking.</li></ul>

**All terms and conditions will remained same.**

**Stores Officer (H)  
AIIMS Raipur (C.G)**