



Date: - 27/05/2019

Corrigendum

विषय / Sub: Corrigendum in Notice Inviting Quotation for “Supply of consumable items i.e. Genotyping Assay” for Genetics Department at AIIMS, Raipur

संदर्भ / Ref: AIIMS/R/CS/03/GENETICS/19/LPC

In context to following corrigendum is issued against subject LPC:

Sr. No	Item Detail (Existing)	To be read as
	LPC opening Date: 29.05.2019, 3:30P.M.	LPC opening Date: 03.06.2019, 3:30P.M.
01	SNP Genotyping Assay, for MTHFR gene A1298C or E429A (RS1801131)	SNP Genotyping Assay, for MTHFR gene A1298C or E429A (RS1801131) : Using TaqMan, single-tube format, pre-designed to provide highly accurate, reproducible and reliable results. Each pre-formulated assay should contain: 2 unlabelled PCR primers: (forward and reverse (primers at 900 nM final concentration), VIC dye—MGB labeled probe to detect the Allele 1 sequence (probes at 200 nM final concentration) and 01 6FAM dye—MGB labeled probe to detect the Allele 2 sequence (probes at 200 nM final concentration) compatible to run on Biorad CFX96 Realtime Machines. sufficient to run 300 reactions
02	SNP Genotyping Assay, for BRCA1 185delAG (RS386833395)	SNP Genotyping Assay, for BRCA1 185delAG (RS386833395) : Using TaqMan, single-tube format, pre-designed to provide highly accurate, reproducible and reliable results. Each pre-formulated assay should contain: 2 unlabelled PCR primers: (forward and reverse (primers at 900 nM final concentration), VIC dye—MGB labeled probe to detect the Allele 1 sequence (probes at 200 nM final concentration) and 01 6FAM dye—MGB labeled probe to detect the Allele 2 sequence (probes at 200 nM final concentration) compatible to run on Biorad CFX96 Realtime Machines. sufficient to run 300 reactions
03	SNOP Genotyping Assay, for BRCA1 5382insC (RS80357906)	SNOP Genotyping Assay, for BRCA1 5382insC (RS80357906) : Using TaqMan, single-tube format, pre-designed to provide highly accurate, reproducible and reliable results. Each pre-formulated assay should contain: 2 unlabelled PCR primers: (forward and reverse (primers at 900 nM final concentration), VIC dye—MGB labeled probe to detect the Allele 1 sequence (probes at 200 nM final concentration) and 01 6FAM dye—MGB labeled probe to detect the Allele 2 sequence (probes at 200 nM final concentration) compatible to run on Biorad CFX96 Realtime Machines. sufficient to run 300 reactions
04	SNP Genotyping Assay, for BRCA2 617delT (RS80359550)	SNP Genotyping Assay, for BRCA2 617delT (RS80359550) : Using TaqMan, single-tube format, pre-designed to provide highly accurate, reproducible and reliable results. Each pre-formulated assay should contain: 2 unlabelled PCR primers: (forward and reverse (primers at 900 nM final concentration), VIC dye—MGB labeled probe to detect the Allele 1 sequence (probes at 200 nM final concentration) and 01 6FAM dye—MGB labeled probe to detect the Allele 2 sequence (probes at 200 nM final concentration) compatible to run on Biorad CFX96 Realtime Machines. sufficient to run 300 reactions

05	SNP Genotyping Assay, for BRAF V600E (RS113488022)	SNP Genotyping Assay, for BRAF V600E (RS113488022) : Using TaqMan, single-tube format, pre-designed to provide highly accurate, reproducible and reliable results. Each pre-formulated assay should contain: 2 unlabelled PCR primers: (forward and reverse (primers at 900 nM final concentration), VIC dye—MGB labeled probe to detect the Allele 1 sequence (probes at 200 nM final concentration) and 01 6FAM dye—MGB labeled probe to detect the Allele 2 sequence (probes at 200 nM final concentration) compatible to run on Biorad CFX96 Realtime Machines. sufficient to run 300 reactions
----	---	--

Note: All other terms & condition will be remain unchanged.

Stores Officer
AIIMS Raipur (CG)