



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)  
All India Institute of Medical Sciences, Raipur (Chhattisgarh)  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

AIIMS/R/2017/HS/A.A1 block/8218/LPC155

Date: 12-05-2017

**Inviting Quotation for Purchase of medical equipments for A.A1 block in hospital at AIIMS Raipur.**

**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of medical equipments for A.A1 block in hospital at AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Store Officer Room No. – 51, 2nd floor Ayush Buliding, Gate No. 1 up to 23-05-2017 before 3:00 pm.** The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

**Hospital Requirement of medical equipments for A.A1 block in hospital at AIIMS Raipur. .**

Sr No.	Complete Description of items (Specification Model, Catalog No) Use separate Sheet if required & signed by indenter and HOD	Quantity required	Unit	Brand /Make	Unit rate inclusive taxes (INR)	Amount (INR)
A	<b>Specification for silicone adult resuscitator :</b>  1. Semi-transparent Resuscitator bag adult with face mask of size 5 deliver a max. Tidal volume of approximate 1500 ml , the outer cover of bag should be 100% latex free , with Single shutter Patient Valve. The bag should be made of silicone rubber. 2. It should have expiratory connector (for PEEP valve attachment): 30 mm male (ISO). 3. It should have hand strap ensures a good grip, which helps to reduce fatigue during manual ventilation. It should have single shutter valve. It should have double swivel mount at mask connector and bladder enable free moment of hands	08	Nos			

	<p>without disrupting manual ventilation. This valve impacts each stroke and retains oxygen level within reservoir bag.</p> <p>4. Volume of oxygen reservoir bag is approx. 1500 ml.</p> <p>5. Resuscitators can be autoclaved repeatedly at 134°C.</p> <p>6. It should be CE/ISO, US FDA certified.</p>					
<b>B</b>	<p><b>Specification for silicone infant/child resuscitator:</b></p> <p>1. Semi-transparent Resuscitator bag Infant / Child with face mask of size 2 deliver a max. tidal volume of approximate 450 ml , the outer cover of bag should be 100% latex free , with Single shutter Patient Valve. The bag should be made of silicone rubber.</p> <p>2. It should have expiratory connector (for PEEP valve attachment): 30 mm male (ISO).</p> <p>3. Volume of oxygen reservoir bag is approx. 1500 ml.</p> <p>4. It should have hand strap ensures a good grip, which helps to reduce fatigue during manual ventilation. It should have single shutter valve. It should have double swivel mount at mask connector and bladder enable free moment of hands without disrupting manual ventilation. This valve impacts each stroke and retains oxygen level within reservoir bag.</p> <p>5. Resuscitators can be autoclaved repeatedly at 134°C.</p> <p>6. It should be CE/ISO,US FDA certified.</p>	08	Nos			

#### **Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. Tin number and relevant document should be submitted along with quotation.
4. **Warranty of the item should be at least one year.**

5. Products are certified from ISI//FDA/CE0434/, the Certificate to this effect should be attached.
6. Supply should be done within 15 days after Placement of PO.
7. Price should be FOR Destination basis (i.e. concerned department).
8. Payment will be released after certification from Representative member of Medical Superintendent.
- 9. Quotation Name/No. and due date of opening must be mentioned on top of envelopes.**
10. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
11. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
12. All other terms & condition as per GCC applicable.
13. Material to be delivered at **Gate No-01, Hospital Store, AYUSH PMR Building, 2<sup>nd</sup> Floor, Room No.-56.**
14. **Validity of the quotation should be 90 days from the date of opening.**
15. **Sample to be submitted as and when Required.**

**Rishi Gupta**  
**Stores Officer (H)**  
AIIMS, Raipur (C.G.)