



AIIMS/R/CS /Patho/17/018/LPC

Date: 15.06.2017

Inviting Quotations for Purchase of **STIRRED/CIRCULATING WATER BATH** for use in hematology lab for coagulation test Department of Pathology & Lab Medicine, AIIMS Raipur.

**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of **STIRRED/CIRCULATING WATER BATH** Department of Pathology & Lab Medicine, AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Store Officer Medical College Building** up to 23/06/2017 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

S.No.	Particulars	Req. Qty	Per Unit Cost
1	<b><u>STIRRED/CIRCULATING WATER BATH</u></b>	01	

**Technical Specification for Stirred/Circulating Water Bath:**

1. Temperature range (metric) of the septum should be able to maintain Ambient. + 5<sup>0</sup> C to 100<sup>0</sup> C
2. Should have heating Capacity up to 1200 W or above
3. Bath Volume should be approx. 19L(5.0 GAL)
4. Amperage should be 9.0-10.5/4.3-5.0 (120V/230V)
5. Should have Monochrome LCD Display
6. Temperature Should be able to maintain Uniformity of  $\pm 0.05$  °C
7. Cabinet Material should be of Epoxy Powder Coated Cold Rolled Steel
8. Chamber Material should be made up of 304 Stainless Steel
9. Company should have UL & CE Certifications/ Compliance
10. Dimensions (L x W x H) Exterior should be around 16x25x10 in approx.
11. Electrical Requirements-Global Voltage 100-115V/200-230V, 50/60Hz
12. Should be digital controller type
13. Temperature Stability should be  $\pm 0.1$  °C
14. Should have 4 Temperature Presets
15. 2 Stainless Steel Gable Covers, 2 Diffuser Trays, and Rubber Duck should be supplied with this instrument
16. Should achieve temperature uniformity with perimeter – directed water flow.
17. Should be easy to clean and maintain, bath should be coil-free internal design
18. Should have optimize scheduling with auto-on and auto-off timers
19. Should accommodate taller lab- ware with hinged lid and extended height
20. Should help prevent bath overheating and damage with low-fluid protection
21. Should be easy to operate and monitor with icon- based graphical display
22. Should protect the work with audible alarms.
23. Should be supplied with calibration certificate.
24. Should be supplied with 2 years warrantee.

### **OTHER COMMERCIAL TERMS**

1. Rate should be mentioned in words & figure both
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. VAT Reg. (i.e. Form 11) must be attached with the quotation.
8. 100% payment against receipt and acceptance of material.
9. RTGS detail required for payment purpose.
10. Brand & Make should be clearly mentioned in offer.
11. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.
12. Validity of offer should not be less than 90 days

**Store Officer  
AIIMS, Raipur (C.G.)**