# aughary 7

### अखिल भारतीय आयुर्विज्ञान संस्थान, रायप्र(छतीसगढ़)

## All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

Date: 24.06.2017

#### AIIMS/R/CS /Patho/17/018/LPC

Inviting Quotations for Purchase of <u>STIRRED/CIRCULATING WATER BATH</u> for use in hematology lab for coagulation test Department of Pathology & Lab Medicine, AIIMS Raipur.

#### **QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of **STIRRED/CIRCULATING WATER BATH** Department of Pathology & Lab Medicine, AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Store Officer Medical College Building** up to 30/06/2017 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

S.No.	Particulars	Req. Qty	Per Unit Cost
1	STIRRED/CIRCULATING WATER BATH	01	

#### **Technical Specification for Stirred/Circulating Water Bath:**

- 1. Temperature range (metric) of the septum should be able to maintain Ambient. + 5° C to 100° C
- 2. Should have heating Capacity up to 1200 W or above
- 3. Bath Volume should be approx. 19L(5.0 GAL)
- 4. Amperage should be 9.0-10.5/4.3-5.0 (120V/230V)
- 5. Should have Monochrome LCD Display
- 6. Temperature Should be able to maintain Uniformity of  $\pm$  0.05 °C
- 7. Cabinet Material should be of Epoxy Powder Coated Cold Rolled Steel
- 8. Chamber Material should be made up of 304 Stainless Steel
- 9. Company should have UL & CE Certifications/ Compliance
- 10. Dimensions (L x W x H) Exterior should be around 16x25x10 in approx.
- 11. Electrical Requirements-Global Voltage 100-115V/200-230V, 50/60Hz
- 12. Should be digital controller type
- 13. Temperature Stability should be ±0.1 °C
- 14. Should have 4 Temperature Presets
- 15. 2 Stainless Steel Gable Covers, 2 Diffuser Trays, and Rubber Duck should be supplied with this instrument
- 16. Should achieve temperature uniformity with perimeter directed water flow.
- 17. Should be easy to clean and maintain, bath should be coil-free internal design
- 18. Should have optimize scheduling with auto-on and auto-off timers
- 19. Should accommodate taller lab- ware with hinged lid and extended height
- 20. Should help prevent bath overheating and damage with low-fluid protection
- 21. Should be easy to operate and monitor with icon-based graphical display
- 22. Should protect the work with audible alarms.
- 23. Should be supplied with calibration certificate.
- 24. Should be supplied with 2 years warrantee.

#### OTHER COMMERCIAL TERMS

- 1. Rate should be mentioned in words & figure both
- 2. Taxes, if any (should be clearly mention).
- 3. Delivery Schedule within 10 days from the date of issue of PO.
- 4. Price should be FOR Destination basis.(i.e. concerned department)
- 5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
- 6. Quotation No/Name and Due date of opening must be written on top of envelop.
- 7. VAT Reg. (i.e. Form 11) must be attached with the quotation.
- 8. 100% payment against receipt and acceptance of material.
- 9. RTGS detail required for payment purpose.
- 10. Brand & Make should be clearly mentioned in offer.
- 11. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.
- 12. Validity of offer should not be less than 90 days

Store Officer AIIMS, Raipur (C.G.)