



क्रमांक /No. AIIMS/R/CS/T&E/17/23-276/LPC

दिनांक: 29/09/2017

Notice for inviting Quotation for supply of Activated Clotting Time Machine for Trauma & Emergency, at AIIMS Raipur.

कोटेशन सूचना
QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having TIN/ relevant documents for supply of the following items to be used in Trauma & Emergency at AIIMS, Raipur and should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on **06th October 2017**, item description as per detailed bellow.

क्र.स. S. N.	सामग्री का विवरण Description of Item	एचएसएन कोड HSN Code	मात्रा Qty.	इकाई दर Unit Price in ₹	कर GST in %	कर सहित इकाई दर Unit Price Incl. Tax.	कुल रकम Total Price in ₹.
1.	<u>ACT Machine</u> Specification of ACT Machine: ACT Machine Should be capable to measure whole Blood Activated Clotting time (ACT) with celite, Kaolin, glass-bead activators; ACT Tubes with blended activators. The ACT Machine should be portable and well suited for all point-of-care testing situations. 1. 2 Point clot detection facility should be available to get accurate results. 2. Single key operation should be available 3. Data management and connectivity-ready capabilities should be available. 4. It should be microprocessor-controlled electromechanical coagulation instrument designed to determine coagulation endpoints in fresh and citrated whole blood samples. 5. Clotting-time results should be presented on the instrument's large size Liquid Crystal Display (LCD) screens.		01 No				

	<p>6. It should provide long-term data storage and retrieval, with access to previous patient and quality control test data.</p> <p>7. Test results should be stored in the order in which they are performed. The ability to externally transfer or save data enhances the system's versatility.</p> <p>8. All Technology should be mandatorily with safety approval – FDA & CE approved. Approval certificate is to be submitted for the total unit.</p> <p>9. Manufacturer/Supplier should have ISO certification for quality standards.</p> <p>10. Shall comply with IEC 60601-2-16 SAFETY requirements of medical electric equipment.</p> <p>11. Should be supplied with 100 Nos. ACT tube</p> <p>12. It should have warranty of 1 Year.</p>						
--	---	--	--	--	--	--	--

नियम व शर्तें

Terms & Condition

1. Rate should be mentioned in word and figures both.
2. GST (Kindly mention in above table) should be clearly mentioned in the offer.
3. Delivery Schedule – within 15 days from the date of issue of PO.
4. Price should be F.O.R. Destination basis (Trauma & Emergency).
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Quotation No/Name and Due date of opening must be written on top of the envelop.
7. 100% payment against received and acceptance of material.
8. Brand and make should be clearly mentioned in offer.
9. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
10. RTGS detail required for payment purpose.
11. Validity of offer should not be less than 90 days.
12. Supply, Installation and commissioning will be done by firm (if applicable).
13. Firm to submit documentary evidence in support of claim for GST at the time of submission of Bills.

(सुशील सोनबेर)
भंडार अधिकारी
एम्स रायपुर (छ.ग.)