



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/R/CS /Pharma/17/008/LPC

Date: 14.11.2017

कोटेशन सूचना
QUOTATION NOTICE

Inviting Quotations for Purchase of **Instruments** for Clinical and PG Research laboratories for Department Pharmacology, AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Instruments** for Clinical and PG Research laboratories for Department Pharmacology, AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer Medical College Building** up to 24/11/2017 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

क्र. सं. S.No.	सामग्री का विवरण Description of Items	एचएसएन कोड HSN Code	मेक एवं ब्रांड Make / Brand	मात्रा Qty	इकाई दर रु. में Unit Rate in Rs.	टेक्स GST	कर के साथ दर Unit Rate with TAX	कुल मूल्य Total Price
1	Digital hand Steadiness Tester-Hole Type (Electrical)			01 Nos.				
2	Digital Hand Steadiness Tester (Liner Type)			01 Nos.				

Sr. No.	Equipment	Detail
1	Digital Hand Steadiness Tester-Hole Type (Electrical)	<p>Purpose: To measure the psycho-motor functions and visual-motor coordination of a human subject</p> <p>Specification:</p> <ol style="list-style-type: none"> 1. The apparatus should consist of a metallic rust proof panel with a set of 9 or more Circular Holes arranged in decreasing order of diameter from approximately 20 to 2mm 2. A probe or stylus, having a tip of approximately 1.5 mm in diameter and a connecting wire 3. A four figure Electronic Digital Timer with start, reset and stop functions 4. Signal indicator: A beep and light to indicate touching of stylus to the side of a hole 5. Score counter should display both correct score and number of faults or number of holes successfully completed 6. Power ON - OFF Switch should be provided 7. The entire unit should be housed in a sturdy metal enclosure made of good quality rust and corrosion proof materials and duly painted. 8. Operating Manual should be provided 9. Minimum 2 years warranty 10. Operating Power: 220volt, $\pm 10\%$, 50 Hz., A.C, with plug compatible with local sockets. 11. The manufacturer should be ISO certified. The product should have ISI mark.

Sample Image:



Sr. No.	Equipment	Detail
2	Digital Hand Steadiness Tester (Linear Type)	<p>Purpose: To measure the psycho-motor functions and visual-motor coordination of a human subject</p> <p>Specification:</p> <ol style="list-style-type: none"> 1. Should have a curvi-linear groove cut out on a metallic rust proof panel 2. A stylus to run through the groove, with a connecting wire 3. Signal indicator: A beep and light to indicate touching of stylus to the edge of the groove. 4. Should be able to measure distance of movement of stylus. 5. Digital Timer with start, reset and stop functions 6. Digital Error counter 7. The entire unit should be housed in a sturdy metal enclosure made of good quality rust and corrosion proof materials and duly painted 8. Operating Manual should be provided 9. Minimum 2 years warranty 10. Operating Power: 230 V \pm 10% AC, 50 HZ, with plug compatible with local sockets. 11. The manufacturer should be ISO certified. The product should be ISI marked.

Sample Image:



OTHER COMMERCIAL TERMS

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
9. the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
11. The GST registration details may please be furnished.
12. 100% payment against receipt and acceptance of material.
13. Validity of offer should not be less than 90 days
14. No Part supply or Part Payment will be entertained.
15. RTGS detail required for payment purpose.
16. Expenditure will be debitable to GIA-48.
17. Brand & Make should be clearly mentioned in offer (If require).
18. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
19. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.
20. Warranty mention on specification

(सुशील सोनबेर)

भंडार अधिकारी

अखिल भारतीय आयुर्विज्ञान संस्थान

रायपुर (छ.ग.)