



क्रमांक: AIIMS/R/CS/CTVS/17/68/LPC

दिनांक: 30-11-2017

कोटेशन सूचना
QUOTATION NOTICE

Inviting quotations for purchase of “**Activated Clotting Time (ACT) Machine**” for Department of Cardiovascular & Thoracic Surgery (CTVS) at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest/Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2nd Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **06-12-2017** before **3:00** pm. The quotations will be opened on the same day at **3.30 pm**. Details of item are given as under:

क्र स S. No	सामग्री का विवरण Description of item(s)	मात्रा Qty.	HSN Code	ब्रांड Make/ Brand/ Model	ईकाई दर Unit Price in ₹.	कर Tax/ GST	कुल ईकाई रकम (कर सहित) Total Unit Price with Tax in ₹.	कुल रकम Total Amount in ₹.
1.	Activated Clotting Time (ACT) Machine: <u>विनिर्देश (Specification):</u> Detail technical specification as per Annexure-I	01 No.						

नियम एवं शर्तें :

Terms & Condition:

1. Rate should be mentioned in words & figure both.
2. Taxes/GST, if any (Kindly mention in above table)
3. Delivery schedule: within 15 days from the date of issue of P.O.
4. No any additional documents related to this NIQ will be entertained after NIQ opened.
5. Price should be FOR Destination basis (i.e. Department of Cardiovascular & Thoracic Surgery, AIIMS Raipur).
6. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
7. Quotation No/Name and due date of opening must be written on top of envelop.
8. Brand, Make & Warranty should be clearly mentioned in offer.
9. GST rates applicable on your quoted item may please be confirmed. HSN code for each item should be clearly mentioned.



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

10. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
11. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
12. The GST registration details may please be furnished.
13. RTGS details required for payment purpose.
14. 100% payment against receipt and acceptance of material.
15. No part supply or part payment will be entertained.
16. Validity of offer should not be less than 90 days.
17. Supply, installation and commissioning will be done by firm (if applicable).
18. The quantity in above column is totally tentative. It can be increase or decrease at the time of placement of order.
19. AIIMS, Raipur reserve the right to place the order for full or full or part quantity to one or more firms.

(सुशील सोनबेर)
भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपूर (छ.ग.)



ANNEXURE-I

Specification of ACT Machine:

ACT Machine Should be capable to measure whole Blood Activated Clotting time (ACT) with celite, Kaolin, glass-bead activators; ACT Tubes with blended activators. The ACT Machine should be portable and well suited for all point-of-care testing situations.

1. 2 Point clot detection facility should be available to get accurate results.
2. Measurement range 0-1500 sec
3. Incubation temperature 36.5 C – 37.5 C
4. Single key operation should be available
5. Data management and connectivity-ready capabilities should be available.
6. It should be microprocessor-controlled electromechanical coagulation instrument designed to determine coagulation endpoints in fresh and citrated whole blood samples.
7. Clotting-time results should be presented on the instrument's large size Liquid Crystal Display (LCD) screens (8-2-x2.5-3cm).
8. It should provide long-term data storage and retrieval, with access to previous patient and quality control test data.
9. Test results should be stored in the order in which they are performed. The ability to externally transfer or save data enhances the system's versatility.
10. All Technology should be mandatorily with safety approval – FDA approved.
11. Should be supplied with 100 Nos. ACT tube.
12. It should have warranty of 1 year.

(सुशील सोनबेर)
भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपूर (छ.ग.)