



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/R/H.S/2018/SAP-28/33/155

Date: 12-02-2018

Inviting Quotations for Purchase of Sterilization Indicator for autoclave for Hospital Infection control department, at AIIMS Raipur. Under Swachhta Actipn Plan.

कोटेशन सूचना

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Sterilization Indicator for autoclave for Hospital Infection control department, at AIIMS Raipur, Under SAP**. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer Room No. - 51, 2nd floor Ayush Building, Gate No. 1 up to 19-02-2018 before 3:00 pm**. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

Sterilization Indicator for autoclave for Hospital Infection control department, at AIIMS Raipur

| Sr No क्र. | Name of Item समग्री विवरण | Quantity required आवश्यक मात्रा | Unit इकाई | Brand/ Ma ke नि र्मा ण | Produc t HSN code प्रोडक्ट HSN कोड | Unit rate (INR) इकाई दर | Applicable GST (INR) लागु GST | | | Total Amount (INR) कुल दर |
|------------------|---|--|--------------|---------------------------------------|---|-------------------------------------|----------------------------------|------|------|------------------------------------|
| | | | | | | | CGST | SGST | IGST | |
| 1 | Sterilization Indicator for autoclave CLASS-1 (autoclave roll) ➤ It is lead free ➤ It is Heat sensitive chemical indicator with self adherence tape. ➤ SIt does not leave any adhesive residue for stream sterilization cycle. ➤ Uniform color change from white to dark brown for easy | 1Roll ×5 | Roll | | | | | | | |

| | | | | | | | | | | |
|---|---|--|---------------------------|--|--|--|--|--|--|--|
| | <p>identification after after processed in Steam autoclave of different time and temperature requirement , with in 15 minutes at 121 degree C and 3minutes at 134 degree C. this confirms to ANSI/AAMI/ISO 11140-1: 2005.it has FDA clearance.</p> | | | | | | | | | |
| 2 | <p>Sterilization Indicator for autoclave CLASS-II (Bowie – Dick Porous Test Pack)</p> <ul style="list-style-type: none"> ➤ This test pack is for monitoring vacuum formation in Autoclave chamber. The test sheet along with penetration challenge Material is in a single easy handled pack. ➤ It is lead free ➤ The weight per pack is between 110 to 125 gms and compliant to ANSI/ AAMI/IOS 11140-1: 20051. And also compliant to ANSI/AAMI/ISO 11140-5: 20072. It has 510 (k) registration. | <p>1 Box (30 PCs) × 2</p> | <p>Box</p> | | | | | | | |
| 3 | <p>Sterilization Indicator for autoclave CLASS -V (Integrating indicator)</p> <ul style="list-style-type: none"> ➤ The Moving front chemical integrators to verifies all perimeters of steam sterilization together i.e. Time, Temperature and Quality of steam. ➤ The read out indicates accept or reject after processing. | <p>1 packet (500 pcs) × 2</p> | <p>pac ket</p> | | | | | | | |

| | | | | | | | | | | |
|--|---|--|--|--|--|--|--|--|--|--|
| | <p>➤ The chemical is inside a lamination to ensure any chemical transfer from indicator to pack/instrument. This is compliant to BS/EN/ISO 11140-: 2005 it also has should have 510 (k) registration.</p> | | | | | | | | | |
|--|---|--|--|--|--|--|--|--|--|--|

Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST number and relevant document should be submitted along with quotation.
4. Products are certified from ISI/FDA/CE0434/GMP certified as applicable.
5. Supply should be done within 15 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. Payment will be released after certification from **Hospital Infection control department.**
8. **Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity All other terms & condition as per GCC applicable.
11. Material to be delivered at **Gate No-01, Ayush PMR Building Hospital General Store 2nd flore room no 56.**
12. **Validity of the quotation should be 90 days from the date of opening.**
13. **Sample to be submitted as and when required.**
14. **Quotation must be in the given format.**
15. **Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.**

Rishi Gupta
Stores Officer (H)
 AIIMS, Raipur (C.G.)