



AIIMS/R/2018/HS/Laryngoscope/01/LPC/155/

Date 23/02/2018

Inviting Quotations for Purchase of Non Consumable items for the Hospital Store AIIMS Raipur.

**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of Non Consumable for Hospital Store. The quotation with copy of certificate of TIN & other documents should be submitted to office of Store Officer Room No. - 51, 2nd floor Ayush Building, Gate No. 1 up to 01.03.2018 before 3.00 pm. The quotations will be opened on the same day at 3.30pm. Details of item are given as under:-

S. no.	Complete Description of Items	Qty.	Make/Model/Brand	HSN Code	Basic Unit Rate	Applicable GST %			ALL inclusive Cost
						SGST	CCST	IGST	
1	Laryngoscope Adult Specification attach Annexure 1	25							
2	Laryngoscope Pediatric Specification attach Annexure 1	10							
3	Laryngoscope Infant Specification attach Annexure 1	10							

**Terms & Condition**

1. Firm to mention Manufactures name/Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. Tin number and relevant document should be submitted along with quotation.
4. Supply should be done within 15 days after Placement of PO.
5. Price should be FOR Destination basis (i.e. concerned department).
6. Payment will be released after certification from Medical representative AIIMS Raipur.
7. Quotation Name/No. and due date of opening must be mentioned on top of envelopes.
8. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
9. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
10. All other terms & condition as per GFR - 2017 applicable.
11. Material to be delivered at Gate Ayush PMR Building Room No. 56, Hospital Store,
12. Validity of quotation should be 90 days from the date of opening

Stores Officer (H)  
AIIMS, Raipur (C.G.)

## Annexure - 1

Technical specification of pediatric and adult laryngoscope.

1. CE or FDA/ISO certified direct laryngoscope handle with detachable blades.
2. Portable, shall not be bulky.
3. Equipment shall be sturdy, high quality stainless steel, 304, matt/glossy finish, resistant to corrosion, resistant to breakage and damage on fall on hard surface or hit against wall, shall bear drop test of 2 meter.
4. Handle of standard size and pencil light handle. It shall have shell and battery with LED/halogen lamp technology, Battery shall be 2.5 v rechargeable/non rechargeable with a capacity of not less than 2 Hrs. of illumination. Handle and blade shall be interchangeable.
5. Classic Macintosh blade size 1, 2, 3 and 4 made of high quality stainless steel matt/glossy finish, light source with LED/halogen lamp technology. Blade shall be smooth, smooth edges to avoid injury, will be maintenance free.
6. Classic Miller blade size 00,0,1,2 made of high quality stainless steel matt finish, light source with LED/halogen lamp technology. Blade shall be smooth, smooth edges to avoid injury, maintenance free.
7. Light source in oropharynx should be balanced illumination. Temperature of light source or blade should not exceed  $>40^{\circ}\text{c}$  even on prolonged use.
8. Warranty period of the laryngoscope shall be 5 years.

Laryngoscope with blade adult		
Sr. No.	Equipment name	Type
1	Standard handle	
2	Classic Macintosh blade	0
		1
		2
		3
		4
3	Battery	1 set

Laryngoscope with blade pediatrics		
Sr. No.	Equipment name	Type
1	Pen light handle	
2	Classic miller blade	00
		0
		1
		2
3	Battery	1 set