



क्रमांक /No. AIIMS/R/CS/T&E/17/23-361/LPC

दिनांक: 14/03/2018

Notice for inviting Quotation for supply of Suction Machine for Trauma & Emergency, at AIIMS Raipur.

कोटेशन सूचना
QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for supply of the following items to be used in Trauma & Emergency at AIIMS, Raipur and should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on **21st March 2018**, item description as per detailed bellow.

क्र.स. S. N.	सामग्री का विवरण Description of Item	एचएसएन कोड HSN Code	मात्रा Qty.	इकाई दर Unit Price in ₹	कर GST in %	कर सहित इकाई दर Unit Price Incl. Tax.	कुल रकम Total Price in ₹.
1.	<u>Suction Machine</u> Specification of Suction Machine: 1. The company should be ISO certified. 2. The unit shall be CE or FDA or ISO certified unit. 3. Complete M.S. powder coating/fiber, hand switch and foot switch and foot switch for operation easily. 4. Absolutely noiseless. 5. Heavy duty motorized pump-for fast vaccum built up. 6. Over flow safety device over flow check valve and removable filter to protect the pump from moisture. 7. Selected negative adjustment and over flow protective mechanism, safe and reliable. 8. Machine should be with polycarbonate jars (Autoclavable) with over flow protection system with bacterial filter. 9. Operating voltage 180-230 volts Ac. 10. Should be UPS inverter friendly. 11. Max vaccum approx.		05 Nos				

	<p>715±10mm Hg regulable through vaccum control Knob.</p> <p>12. Max vaccum 20lits/min vaccum gauge bourdon type 3” 0-760mm Hg calibration.</p> <p>13. Suction jars of capacity 2000-2500ml made up of polycarbonate.</p> <p>14. Motor ¼ Hp vaccum pump rotatory type overall size L330 X W405 X H730mm (approx.)</p> <p>15. Demonstration and training should be provided by the supplier at the time of installation.</p> <p>16. Warranty should be at least 2 years.</p> <p>17. Additional fuse should provide.</p> <p>18. The machine should be portable on castors/mobile trolley.</p> <p>19. At least 5 spare filter with each machine.</p> <p>20. 1 polycarbonate jar extra with each machine.</p> <p>21. Sample of machine should be provided at the time of quotation submission.</p>						
--	---	--	--	--	--	--	--

नियम एवं शर्तें:

Terms & Condition:

1. Rate should be mentioned in word and figures both.
2. Taxes/GST, if any (Should be clearly mentioned).
3. Delivery Schedule – within 15 days from the date of issue of PO.
4. No any additional documents related to this NIQ will be entertained after opening of NIQ.
5. Price should be F.O.R. Destination basis (Trauma & Emergency Department).
6. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
7. Quotation No/Name and Due date of opening must be written on top of the envelop.
8. 100% payment against received and acceptance of material.
9. **Brand and make should be clearly mentioned in offer as well as tender/quotation specific authorization may be submit with the offer/Bid.**
10. GST rate applicable on your quoted item may please be confirmed. HSN code for each item should be clearly mentioned.
11. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
12. In the event of increase in price, detailed justification and supporting evidence may be submitted.

13. The GST registration details may please be furnished.
14. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
15. RTGS detail required for payment purpose.
16. No part supply of part payment will be entertained.
17. Validity of offer should not be less than 90 days.
18. Supply, installation and commissioning will be done by firm (if applicable).
19. The quantity shown in above requirement column are totally tentative, it can be increase and decrease at the time of placement of purchase order.
20. AIIMS Raipur reserve the right to ask the sample for evaluation before placement of order

(सुशील सोनबेर)
भंडार अधिकारी
एम्स रायपुर (छ.ग.)