



क्रमांक /No. AIIMS/R/CS/BB/18/53/LPC/A

दिनांक: 19/04/2018

Notice for inviting Quotation for Non-consumable items for Transfusion
Medicine & Blood Bank at AIIMS Raipur.

कोटेशन सुचना
QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/
Distributors having GST/ relevant documents for supply non-consumable items for
Transfusion Medicine & Blood Bank at AIIMS, Raipur should be submitted to
Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00
pm on **25th April 2018**, item description as per detailed bellow.

S. N.	Description of Item	Qty.	HSN Code	Make/ Brand/ Model	Unit Price in ₹	GST in %	Unit Price Incl. GST	Total Price in ₹.
1.	Portable Blood Donor Chair विनिर्देश: (Specification): Detail technical specification as per Annexure-I	06 Nos		Remi or Similar				
2.	Tube Stripper विनिर्देश: (Specification): Detail technical specification as per Annexure-I	04 Nos		Terumo Penpol or Similar				
3.	Plasma Expressor (Manual) विनिर्देश: (Specification): Detail technical specification as per Annexure-I	02 Nos		Terumo Penpol or Similar				

नियम व शर्तें

Terms & Condition

1. Rate should be mentioned in word and figures both.
2. GST (Kindly mention in above table) should be clearly mentioned in the offer.

3. Delivery Schedule – within 15 days from the date of issue of PO.
4. Price should be F.O.R. Destination basis (Transfusion Medicine & Blood Bank).
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Quotation No/Name and Due date of opening must be written on top of the envelop.
7. 100% payment against received and acceptance and satisfactory installation of material.
8. Brand and make should be clearly mentioned in offer.
9. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
10. RTGS detail required for payment purpose.
11. Validity of offer should not be less than 90 days.
12. Supply, Installation and commissioning will be done by firm (if applicable).
13. Firm to submit documentary evidence in support of claim for GST at the time of submission of Bills.
14. The quantity shown in above column are totally tentative, it can be increase and decrease at the time of placement of order.
15. The bidder must quote for 1 year onsite warranty for Portable Blood Donor Chair as per the above mentioned Specification from the date of completion of the satisfactory installation as certified by the department.
16. The bidder should submit their quote for subsequent 1 year onsite AMC for Portable Blood Donor Chair as per annexure-B but it should not be more than 4% of the quoted unit price.
17. Failure to comply this condition will entail the rejection of the quotation/bids. The Price comparison shall be made taking into account on basis price and post warranty AMC. The Rate Contracting Authority reserve the right to award AMC shall be quoted for equipment. So the price of AMC should be quoted according to the cost of equipment.
18. The amount of AMC would be released to the supplier on successful completion of the maintenance of that particular year duly certified by the department.
19. Quoting price for Portable Blood Donor Chair and AMC is mandatory. Lowest one will be deciding i.e. **L1 = Item price with 01 year warranty + 01 year AMC**
20. The 10% of total item value may be with held as security deposit during the warranty period of assets (if applicable).

(सुशील सोनबेर)
भंडार अधिकारी
एम्स रायपुर (छ.ग.)

1) Technical Specifications- for Portable Blood Donor Chair

1. Economic Design & Comfort sitting.
2. Stainless steel 304 Anti corrosive structure
3. Seat Width: 600mm, Length: 1700mm,
4. Seat height from floor: 650 – 700mm
5. Gross Weight: 15kg
6. Can Bear Donor Weight up to 120 Kg
7. Adjustable hand rest position
8. Couch easily reclined to syncope secure position.
9. Pocket at the back of each couch for accessories
10. User list for satisfactory performance from three end users.
11. The equipment should be a BIS approved or CE European certified product.

2) Technical Specifications- for Tube Stripper

1. Should have completely anti-rust, stainless steel body.
2. Should be light weight.
3. Should ensure the uniform pressure while pressing to close and automatic recoiling of spring to release handle for opening.
4. Should have Screw-less rollers to avoid loosening of the rollers.
5. Should have extra sharp cutting edges.
6. Should behave ergonomically designed handle for better grip.
7. Should have roller guide to avoid any damage of tube.
8. Should have provision for manual tube sealing buy aluminium rings
9. Original literature of equipment should be submitted
10. Should be BIS certificate: ISO 13485 approved product.
11. One year manufacture warranty.
12. Users list should be attached with satisfactory report the last three years from three users with contact details.

3. Technical Specifications- for Plasma Expressor (Manual)

1. Should be suitable to manually express blood components (Plasma, Platelets) from collection blood bags.
2. Front panel should be spring loaded to apply uniform pressure on container causing transfer of fluid.
3. Compression plate should be made of durable transparent acrylic material.
4. Metal used for the apparatus should be non-corrosive and can be cleaned with antiseptics.
5. Base portion and vertical surface should be made to have better strength and long lasting performance.
6. Certification: Product certification: BIS certification or CE class IIA or ISO 13485 product.
7. One year manufacturer warranty (mandatory)
8. Users list should be attached with satisfactory report the last three years from three users with contact details.
9. Demonstration of performance of equipment is compulsory in nearby area failing to which will be disqualification.

(सुशील सोनबेर)
भंडार अधिकारी
एम्स रायपुर

Annexure: B**PRICE SCHEDULE FOR
ANNUAL MAINTENANCE CONTRACT (A.M.C.)****(RATES SHOULD BE QUOTED IN INDIAN RUPEES ONLY)**

Sl. No.	Name of the item	Qty.	For 1 year AMC after warranty (In Rs.)	GST @ %	Net Amount with GST	Total Amount with GST charges
(1)	(2)		(3)	(4)	(5)	(6)
1.	Portable Blood Donor Chair	06 Nos				

Note:-

1. Firm should quote the rate for AMC for each equipment at relevant column and it should not be more than 4% per year of unit price of the quoted price.
2. Firm should upload this sheet after filling of relevant column as attachment of quotation.
3. In case of discrepancy between unit price and total prices, THE LOWER ONE shall prevail.
4. The cost of Annual Maintenance Contract (AMC) after satisfactory completion of Warranty period which include preventive maintenance including testing & calibration as per technical/ service/ operational, should be quoted along taxes applicable on the date of tender opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
5. AMC charges will be negotiable with respect to lower charges quoted for AMC.
6. The uptime warranty will be 98% on 24 (hrs) x 365 (days) basis or as stated in Technical Specification of the TE document.

Place:

Date

Signature of bidder with Seal