



क्रमांक. AIIMS/R/CS/Neuro/103/18/LPC/A

दिनांक: 23.05.2018

विषय : Inviting Quotations for purchase of Aneurysm Clips and Applicators for Neurosurgery Department at AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST No./relevant documents for supply of **Aneurysm Clips and Applicators** for Neurosurgery Department at AIIMS Raipur and should be submitted to **Medical College Building, 2nd Floor, Gate no. 05** office of Store Officer up to 3:00 pm on 01.06.2018. The quotations will be opened on the same day at 03:30 pm. Item description as per detailed below:

S.n. क्र. स.	Name of the Item आइटम का नाम	Qty मात्रा	Make / Model	HSN code एच.एस.एन कोड	UNIT RATE IN Rs. इकाई दर रु में	GST@% जी.एस.टी @%	Unit rate incl. GST यूनिट दर जी.एस.टी सहित	Total amount in Rs कुल मूल्य
01	Aneurysm Clips and Applicators (Specification as per Annexure-A)	01 set						

L-1* will be decided on total value only, so firms are supposed to quote with total amount but firm should furnish the rate of individual items too.

नियम व शर्तें:

- Rate should be mentioned in words & figure both.
- Taxes, if any (should be clearly mention).
- Delivery Schedule – within 15 days from the date of issue of PO.
- Price should be FOR Destination basis.(i.e. concerned department)
- LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
- Quotation No/Name and Due date of opening must be written on top of envelop.
- GST rates applicable on your quoted item may please be confirmed. **HSN code** for each item should be clearly mention.
- Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. “Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
- In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
- The GST registration details may please be furnished.
- 100% payment against receipt and acceptance of material.
- Validity of offer should not be less than 90 days
- No Part supply or Part Payment will be entertained.
- RTGS detail required for payment purpose.
- Expenditure will be debitable to GIA-capital.
- Brand & Make should be clearly mentioned in offer (If require).
- AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.
- The Quantity of above column is totally tentative. It may be increased or decreased at the time of placement of order.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)

Annexure 1

Technical Specification of Aneurysm Clips and Applicators:-

The instruments should be light weight, finely balanced of non-magnetic surgical grade stainless steel with carbon contents specified of ISO 9001/CE/PIS or equivalent rating.

The dimensions of instrument may vary 5-10% from specification mentioned.

It should be BIS/ISO/European CE/USFDA approved.

All instruments should be manufactured by same manufacturer.

Aneurysm Clips and Clip Applicators:-

1.1 Aneurysm Clip Applicators:-

The clip applicator should be made of high grade stainless steel/Titanium. The Quality of steel/Titanium should comply with the DIN standards.

- The applicator should be compatible with both Titanium and Phynox clips.
- It should be high quality and precision.
- It should be light weight, strong and durable.
- It should be non-magnetic.
- All instruments should be autoclavable.
- The manufacturer should have a direct repair facility available in India.
- All clip applicator should be autoclavable at 134deg.
- It should be supplied with sterilization tray for clip applicator.
- Instruments should meet the International Certification as per DIN EN ISO 13845:2001.
- It should have TUV certificate for manufacturer with details of manufacturing facility.
- The manufacturer should have service facility in India.

Sr N.	Specification		Quantity in No.
1	Yasargil Bayonet Aneurysm clip applicator Straight, Working Length 110 mm	For standard Clip.	1
2	Yasargil Bayonet Aneurysm clip applicator with swivel tips and adjustable jaw (Double Hinged). Working length 90 mm. Should be supplied with adjusting Key	For standard Clip.	1
3	Yasargil Aneurysm clip removing forceps 110 mm	For standard Clip.	1

1.2. Aneurysm Clips:-

- The aneurysm clip should be made of Titanium/Cobalt alloy.
- The clip should be MRI safe.
- The clip should have box lock joint to avoid scissoring affect.
- The blade profile should have pyramid like impression. It should be atraumatic.
- The clip should ensure even pressure on the complete blade surface.
- Each clip should have individual serial no. to ensure complete tracking of manufacturing process, material, closing force, sterilization process.
- The clip should be supplied in sterile pack.
- The package of clip should mention article no, expiry date, closing force.
- The clip should be color coded.
- It should be supplied with sterilization tray with detachable marking showing the clip article no. and clip design.

Aneurysm clips

S N	Specifications		Quantity in No.
1	Yasargil Permanent Standard Aneurysm clip Straight	7mm	1
2	Yasargil Permanent Standard Aneurysm clip Straight	9mm	1
3	Yasargil Permanent Standard Aneurysm clip Straight	11mm	1
4	Yasargil Permanent Standard Aneurysm clip Curved	6.5mm	1
5	Yasargil Permanent Standard Aneurysm clip Curved	8.3mm	1
6	Yasargil Permanent Standard Aneurysm clip Curved	10.2mm	1
7	Yasargil Permanent Standard Aneurysm clip Straight with 3.5mm Fenestration	3/8.1mm	1
8	Yasargil Permanent Standard Aneurysm clip Straight with 3.5mm Fenestration	4/9.1mm	1
9	Yasargil Permanent Standard Aneurysm clip Straight with 3.5mm Fenestration	5/10.1mm	1
10	Yasargil Permanent Standard Aneurysm clip Straight with 5.0mm Fenestration	3/9.8mm	1
11	Yasargil Temporary Standard Aneurysm clip Straight	7mm	1
12	Yasargil Temporary Standard Aneurysm clip Curved	8.3mm	1
13	Yasargil Temporary Standard Aneurysm clip Straight	9mm	1

Sterilization tray:

S N	Specifications	Quantity in No.
1	Aneurysm Clip Sterilization tray to hold clips	1
2	Sterilization tray for Aneurysm clip applicator with lid 1	1

Note: - The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Raipur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.