



क्रमांक: AIIMS/R/CS/Pediatric Sur./18/327/LPC/

दिनांक: 07.08.2018

कोटेशन सूचना

(QUOTATION NOTICE)

Quotation inviting for Purchase of Transport Incubator with auto loading collapsible trolley and transport ventilator for Department of Pediatric Surgery at AIIMS, Raipur.

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for Supply/ Purchase of Transport Incubator with auto loading collapsible trolley and transport ventilator for Department of Pediatric Surgery at AIIMS, Raipur. The vendor should submit the quotation with other relevant documents to office of Stores Officer, at 2nd Floor, Medical College Building, Gate No. 05, Tatibandh, GE Road, AIIMS, Raipur, up to **03:00 pm** date: **13.08.2018**. The quotation will be opened on the same day at **03:30 pm**. Detailed specifications of items are as under:

क्रं. Sl. No.	सामग्री का विवरण Item of Description	मात्रा Qty. Req. (Nos.)	मेक Make	एच. एस. एन. कोड HSN Code	इकाई मूल्य Unit Price (₹)	जी. एस. टी. GST	इकाई दर कर सहित Unit Rate with GST	कुल मूल्य Total Price
01	Transport Incubator with auto loading collapsible trolley and transport ventilator (Specification: As per annexure- I)	01						

नियम एवं शर्तें

Terms and Condition

1. Rate should be mentioned in word and figures both.
2. Taxes, if any (Should be clearly mention)
3. Delivery Schedule – within 15 days from the date of issue of P.O.
4. Price should be F.O.R. destination basis (Department of Pediatric Surgery, AIIMS, Raipur).
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed should be deducted.
6. Quotation No. /Name and Due date of opening must be mentioned on top of envelop.
7. GST rates applicable on your quoted item may please be confirmed. HSN code should be clearly mentioned.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. “Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
9. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
10. The GST registration details may please be furnished.
11. No Part supply or Part Payment will be entertained.
12. 100% payment against receipt and acceptance of material. No claim will be entertained regarding interest on any payment of any circumstance.
13. Where there is statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the supplier at rates as notified from time to time.
14. No payment shall be made for rejected stores. Rejected equipment's must be removed by the supplier within two weeks of the dated of issue of rejection advice at their own cost & replace immediately. In case these are not removed these will auctioned at the risk and responsibility of the suppliers without notice.

15. Brand and make should be clearly mentioned in offer (if required).
16. Installation and commissioning will be done by firm (if applicable).
17. RTGS detail required for payment purpose.
18. AIIMS Raipur reserves the right to place order for full of part quantity to one or more firm.
19. Validity of offer should not be less than 90 days.
20. Three years warranty.
21. The quantity shown in above column are totally tentative, it can be increase and decrease at the time of placement of order.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)

विनिर्देश (Specification):

1. Description of Function Required for Transport of Newborn or Premature Baby.
2. Operational Specifications.
3. Complete Transport system with trolley and Incubator
 - 3.1 Shockproof moulded ABS Plastic Construction, lightweight & Sturdy, pre-stressed tropicalized construction.
 - 3.2 Light Weight Construction with incubator not weighing more than 15 Kgs.
 - 3.3 Transparent canopy with mattress.
 - 3.4 Two access doors, take-out mattress tray with baby restraining straps.
 - 3.5 Optimum Baby Area Size 20 inches x 12 inches.
 - 3.6 Warm air circulation system.
 - 3.7 Skin temperature monitoring and microprocessor based PID control: 25 to 38 C, increments 0.1C.
 - 3.8 Digital displays outside shows skin temperature, set temperature and alarms.
 - 3.9 Provision for 2 Oxygen Cylinders in the collapsible trolley.
 - 3.10 Audiovisual alarms: high/low air temperature, overheat at 38 C, temperature sensor failure, power failure and low battery.
 - 3.11 Construction dismantle able allows frequent washing and disinfection of the incubator.
 - 3.12 Battery and AC supported, Power requirements: 220 V / 50 Hz and internal re-chargeable batteries (automatic recharge) .
 - 3.13 In motion, it should work on 12V DC Car Battery for at least 2 hours.
 - 3.14 Power consumption: 100 W
 - 3.15 Device is produced by ISO 13485:2003 certified manufacturer (Certificate to be submitted, further details see "Technical Provisions").
 - 3.16 Manufacturer must have valid CE/FDA/BIS safety approval certificate for Transport Incubators.
 - 3.17 Device should be tested as per IEC 60601-2-20 – Safety Standard for Transport Baby Incubators.
 - 3.18 Should also be provided with transport ventilator for neonate and auto loading trolley.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)