



क्रमांक /No. AIIMS/R/CS/BB/18/04/LPC/C

दिनांक: 12/09/2018

Notice for inviting Quotation for Horizontal Pre Vacuum Autoclave for Transfusion Medicine & Blood Bank at AIIMS Raipur.

कोटेशन सुचना
QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for supply Horizontal Pre Vacuum Autoclave for Transfusion Medicine & Blood Bank at AIIMS, Raipur should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on **19th September 2018**, item description as per detailed bellow.

स. क्र. S. N.	सामग्री का विवरण Description of Item	मात्रा Qty.	HSN Code	ब्रांड Make/ Brand/ Model	इकाई दर Unit Price in ₹	कर GST in %	कर सहित इकाई दर Unit Price Incl. GST
1.	Horizontal Pre Vacuum Autoclave विनिर्देश: (Specification): Detail technical specification as per Annexure-A	01 No					

नियम व शर्तें

Terms & Condition

1. Rate should be mentioned in word and figures both.
2. GST (Kindly mention in above table) should be clearly mentioned in the offer.
3. Delivery Schedule – within 15 days from the date of issue of PO.
4. Price should be F.O.R. Destination basis (Transfusion Medicine & Blood Bank).
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Quotation No/Name and Due date of opening must be written on top of the envelop.
7. 100% payment against received and acceptance and satisfactory installation of material.
8. Brand and make should be clearly mentioned in offer.

9. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
10. RTGS detail required for payment purpose.
11. Validity of offer should not be less than 90 days.
12. Supply, Installation and commissioning will be done by firm (if applicable).
13. Firm to submit documentary evidence in support of claim for GST at the time of submission of Bills.
14. The quantity shown in above column are totally tentative, it can be increase and decrease at the time of placement of order.
15. The bidder must quote for 1 year onsite warranty for Portable Autoclave as per the above mentioned specification from the date of completion of the satisfactory installation as certified by the department.
16. The bidder should submit their quote for subsequent 1 year onsite AMC for Portable Autoclave, as per annexure-B but it should not be more than 4% of the quoted unit price.
17. Failure to comply this condition will entail the rejection of the quotation/bids. The Price comparison shall be made taking into account on basis price and post warranty AMC. The Rate Contracting Authority reserve the right to award AMC shall be quoted for equipment. So the price of AMC should be quoted according to the cost of equipment.
18. The amount of AMC would be released to the supplier on successful completion of the maintenance of that particular year duly certified by the department.
19. Quoting price for Horizontal Pre Vacuum Autoclave and AMC is mandatory. Lowest one will be deciding i.e. **L1 = Item price with 01 year warranty + 01 year AMC**
20. The 10% of total contract value may be with held as security deposit during the warranty period of assets (if applicable).

(सुशील सोनबेर)
भंडार अधिकारी
एम्स रायपुर (छ.ग.)

Horizontal Pre Vacuum Autoclave

Autoclave to be exclusively designed and used for the treatment/disinfection of biomedical waste.

1. Capacity minimum 100 liters.
2. Sterilization temperature 121 deg C to 138 deg C
3. Compressed air 5.0-7.0 BAR
4. Should include vacuum cycle, rapid vacuum cycle and non- vacuum cycle with or without drying.
5. Tamper- proof control panel with efficient display and record devices for critical parameters such as time, temperature, pressure, date and batch number etc. (Touch sensitive control panel).
6. Door with front loading. Door to have a locking mechanism (Door to have a safety lock to allow opening at the end of cooling cycle only).
7. Durable stainless steel pressure vessel (chamber) which conforms to the pressure equipment directive (PED). All metal parts in the inner surfaces to be made of SS.
8. Multi-program model allowing at-least 5 programs to be stored.
9. To be supplied with load probe and printer.
10. Audible and visible alarms including cycle fault and interruption, sterilization failure, low water level and door unlocked.
11. Should have a communication interface for data transfer to a central data capture system.
12. Should work on 200-250 volts at 50 Hz.
13. Containers for autoclave: Stainless steel containers and pans or auto-clavable polypropylene.
14. Each autoclave shall have graphic or computer recording devices which will automatically and continuously monitor and record dates, time of day, load identification number and operating parameters throughout the entire length of the autoclave cycle.
15. Alarm system
 - a. Independent temperature and pressure monitoring
 - b. Cycle information recovery in the case of power failure or cycle interruption
 - c. Fail indicator- indicates cycle failure or interruption
 - d. Door Indicator- indicates the door is unlocked
16. Certifications:
 - Product certification : BIS The National Standards Body of India/ CE Class II A / US FDA certified
 - Quality Certification: ISO certified
 - Electrical Safety: Equipment meets electrical safety specifications such as that of IEC (class I)/ISO standards
 - For indigenous items: should comply with BIS & CPCB standards

(सुशील सोनबेर)
भंडार अधिकारी
एम्स रायपुर

**PRICE SCHEDULE FOR
ANNUAL MAINTENANCE CONTRACT (A.M.C.)**

(RATES SHOULD BE QUOTED IN INDIAN RUPEES ONLY)

Sl. No.	Name of the item	Qty.	For 1 year AMC after warranty (In Rs.)	GST @ %	Net Amount with GST	Total Amount with GST charges
(1)	(2)		(3)	(4)	(5)	(6)
1.	Horizontal Pre Vacuum Autoclave	01 No				

Note:-

1. Firm should quote the rate for AMC for equipment at relevant column and it should not be more than 4% per year of unit price of the quoted price.
2. Firm should upload this sheet after filling of relevant column as attachment of quotation.
3. In case of discrepancy between unit price and total prices, THE LOWER ONE shall prevail.
4. The cost of Annual Maintenance Contract (AMC) after satisfactory completion of Warranty period which include preventive maintenance including testing & calibration as per technical/ service/ operational, should be quoted along taxes applicable on the date of tender opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
5. AMC charges will be negotiable with respect to lower charges quoted for AMC.
6. The uptime warranty will be 98% on 24 (hrs) x 365 (days) basis or as stated in Technical Specification of the TE document.

Place:
Date

Signature of bidder with Seal