



AIIMS/R/CS /Micro/18/90/LPC

Date:29.01.2019

विषय/Sub:Inviting Quotations for procurement of Hot Air Oven 250 Litres for Microbiology Department, at AIIMS Raipur.

**कोटेशन सूचना**  
**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having GST Number and relevant documents for Hot Air Oven 250 Litres for Microbiology Department for Microbiology Department,at AIIMS Raipur. The quotation should be submitted to Medical College Building,2<sup>nd</sup> Floor,Gate no.05 office of Store Officer up to 3:00pm on 04.02.2019. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

क.स. S.No.	सामग्री का विवरण (Description of Items)	आवश्यक मात्रा Requirement	ब्रांड Make/ Model	HSN Code	इकाई दर ₹ में UNIT Rate in Rs.	GST	कुल रकम Total Amount With GST in Rs.
1.	<b>Hot Air Oven</b> <b>Capacity Approx.250 Litres</b> <b>Specification:</b> 1.2 set point Microprocessor based Digital Temperature Controller 2.Tripled walled chambers with insulation; at least 75mm thick high grade glass-wool should be provided to minimize the heat loss. 3.External: Mild steel sheet 20 SWG thick powder coated 4.Toughened Glass window is to be provided to view inside the Oven 5.Insulated mild steel door with lock 6.Castor wheel to be fitted for easy mobility inside the laboratory. 7.Size of the inner chamber approx. 605 X 910 X 605mm (24" X 36" X 24"); 330 liters approx. 8.Inner chamber and three perforated shelves should be made up of 26 SWG thick stainless steel sheet 304 grade 9.Forced air circulation by quiet air turbine/Fan to ensure uniform temperature 10.SS Finned tubular air heater of 3KW capacity has to be provided 11.Digital display of temperature 12.Temperature Variation +/- 2 °C 13.Temperature Range: 5°C to 250 °C 14.Alarm to indicate error or to if chamber temperature exceeds set point	1 No					

15.Power input to be 220-240VAC, 50Hz fitted with Indian plug 16.Voltage corrector/stabilizer of appropriate ratings meeting ISI Specifications.( Input 160-260 V and output 220-240 V and 50 Hz) 17.System should confirm to IS: 6365-1971(Reaffirmed 1995) with latest amendments in ISI specifications for Laboratory Electric Ovens. Alternatively System should be FDA Approved or CE/BIS Certified. 18.Electrical safety conforms to standards for electrical safety IEC-60601 / IS-3450. Should be compliant to ISO 13485: Quality systems - Medical devices - particular requirements for the application of ISO 9001 applicable to manufacturers and service providers that perform their own design activities.						
--	--	--	--	--	--	--

#### नियम एवं शर्तें

#### Terms & Conditions.

1. Rate should be mentioned in words & figures both.
2. **GST rates applicable** on your quoted item may please be confirmed.
3. Delivery Schedule- within 15 days from the date of issue of P.O.
4. No additional documents related to this NIQ will be entertained after opening of NIQ.
5. Price should be F.O.R. for Destination basis (i.e.Concerned Department)
6. LD@0.5% of delayed supply per week or part week for delay of supply of material subject to maximum upto 10% of delayed supply should be deducted.
7. Quotation No/Name and Due date of opening must be written on top of the envelop.
8. **Brand & Make & warranty should be clearly mentioned in offers as well as tender/quotation specific authorization may be submit with the offer/bid.**
9. **The GST registration details may please be furnished.**
10. Please confirm if there any change (Upward/Reduction) in your **Basic Price** structure. And you are also requested to pass the Input Credit as per the following **Anti Profiteering Clause** of GST.**“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”**
11. RTGS details required for payment purpose.
12. 100% payment against receipt and acceptance of material.
13. No part supply or part payment will be entertained.
14. Validity of offer should not be less than 90 days.
15. Supply,Installation and Commissioning will be done by firm (if applicable).
16. The quantity shown in above column are totally tentative, it can be increase and decrease at the time of placement of purchase order.
17. AIIMS Raipur reserved the rights to place order for full or part quantity to one or more firm.

भंडार अधिकारी  
अखिल भारतीय आयुर्विज्ञान  
संस्थान,रायपुर (छ.ग.)

**Vendor Details**

Name	
Aadhaar No. (if any)	
PAN	
GST.NO.	
Address	
City	
State	
Pincode	
Mobile No.	
Phone No.	
E-mail	
Bank Name	
Bank A/c No.	