



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़
(All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

No. AIIMS/R/CS/104/PHY/19/LPC

Dated: 02/02/2019

Notice for inviting Quotation for supply of **Non Consumable items i.e. Centrifuge Machine for Department of Physiology, AIIMS, Raipur.**

कोटेशन सूचना

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST No./relevant documents for supply of Non Consumable item i.e. Centrifuge Machine for Department of Department of Physiology to be used in AIIMS, Raipur and should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on 08.02.2019 . The quotations will be opened on the same day at 3:30pm, item description is detailed below:

S.no.	Item Detail	Qty. Req.	Make	Unit	HSN Code	Unit Price	GST @	Unit Price with Tax	Total Amount
1	Centrifuge Machine <ul style="list-style-type: none">• Should be table top version with angled rotor head• Should have a minimum tube capacity of 24 x15ml.• Should be made up of strong fabricated & corrosion resistant steel and have stainless steel centrifuge, chamber, easy to clean.• Should have control panel-for start/stop switch, dynamic brakes, step less speed regulator with zero start switch protective fuses.• Should have smooth & soft start. Inverter fault detection with auto shutdown.• Should have safety lid interlock to prevent lid opening during centrifugation & automatic door opening on completion.• Should have maintenance free brushless induction motor with frequency drive.• Should have microprocessor controller with digital display.• Max speed range: RPM-6000, accuracy 1RPM, RCF (g)-5070.10.• Power Supply: power input:220-240v AC,50HzS• Should have 7 segment LED display of speed, digital countdown timer (digital timer range 0-59 min) & continuous run.• Should have last set parameters recall, useful for repetitive analysis.• Should have imbalance detection & centrifugation stop with display of error	01							

<ul style="list-style-type: none"> • Should have motor overload protection. • Should have has hinge to prevent door falling. • Should have emergency lid lock release. • Dimension should be approximately (W* D *H):380x470x300mm • Should have ISO 9001:2008 & ISO 13485:2003 certification. • Should have WHO, Good manufacturing practices & CE certification. • Calibration should be performed on supply of instruments & calibration certificate provided. • Company should provide 2 years warranty. 								
---	--	--	--	--	--	--	--	--

नियम और शर्तः

Terms and Condition:

1. Rate should be mentioned in word and figures both.
2. Delivery Schedule – within 15 days from the date of issue of PO.
3. Price should be F.O.R. Destination basis (Department of Physiology, AIIMS Raipur).
4. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
5. Quotation No./Name and Due date of opening must be written on top of envelop.
6. GST rates applicable on your quoted item may please be confirmed. HSN code should be clearly mentioned.
7. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.

“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
8. The GST registration details may please be furnished.
9. RTGS details required for payment purpose.
10. 100% payment against received and acceptance of material.
11. No part supply or part payment will be entertained.
12. No claim will be entertained regarding interest on any payment of any circumstance.
13. Supply, installation and commissioning will be done by firm (if applicable).
14. Where there is statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the supplier at rates as notified from time to time.
15. No payment shall be made for rejected stores. Rejected equipment's must be removed by the supplier within two weeks of the dated of issue of rejection advice at their own cost & replace immediately. In case these are not removed these will auctioned at the risk and responsibility of the suppliers without notice.
16. **Brand and Make should be clearly mentioned in offer as well as tender/quotation specific authorization may be submit with the offer/Bid.**
17. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
18. Validity of offer should not be less than 90 days.
19. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)