



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

AIIMS/R/H.S/2018/ Anaesth/143/155

Date: 28-02-2019

**Inviting Quotations for Purchase of Silicon infant resuscitator (Ambu beg) for Anaesthesia Department Major O.T, AIIMS Raipur.**

**कोटेशन सुचना**

**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Silicon infant resuscitator (Ambu beg) for Anaesthesia Department Major O.T, AIIMS Raipur**. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer Room No. – 51, 2nd floor Ayush Building, Gate No. 1 up to 05-03-2019 before 3:00 pm**. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

**Requirement of Silicon infant resuscitator (Ambu beg) for Anaesthesia Department, Major O.T AIIMS, Raipur**

Sr No क्र.	Name of Item समग्री विवरण	Quantity required आवश्यक मात्रा	Unit इकाई	Brand/Make निर्माण	Product HSN code प्रोडक्ट HSN कोड	Unit rate (INR) इकाई दर	Applicable GST (INR) लागु GST			Total Amount (INR) कुल दर
							C GST	SGST	IGST	
1	<b>Silicon infant resuscitator (Ambu beg)</b> Specifiacion:- 1 Semi –transparent resuscitator bag infant /Child with face mask of size 0 deliver a max tidal volume of approximate 450 ml the outer cover of bag should be 100% latex free with single shutter patient value. The bag should be made of silicon rubber. 2. it should have expiratory connector (for PEEP valve attached ); 30 mm male (ISO) 3. Volume of oxygen reservoir	40	Nos							

	<p>bag is approx 1500 ml</p> <p>4. it should have hand strap ensure a good grip, which helps to reduce fatigue during manual ventilation . it should have single shutter valve. It should have double swivel mount at mask connector and bladder enable free moment of hands without disrupting manual ventilation. This valve impacts each stroke and retains oxygen level within reservoir bag.</p> <p>5. Resuscitators can be autoclaved repeatedly at 134 ° C</p> <p>6. It should be CE/ISO/US FDA certified</p>								
--	--	--	--	--	--	--	--	--	--

### Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST number and relevant document should be submitted along with quotation.
4. Products are certified from ISI//FDA/CE0434/GMP certified as applicable.
5. Supply should be done within 15 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. Payment will be released after certification from Anesthesia Department.
- 8. Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity All other terms & condition as per GCC applicable.
- 11. Material to be delivered at Gate No-04, Hospital Pharmacy Store D-D1 block Lower ground floor.**
- 12. Validity of the quotation should be 90 days from the date of opening.**
- 13. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.**

**Rishi Gupta**  
**Stores Officer (H)**  
 AIIMS, Raipur (C.G.)

