



क्र./ AIIMS/R/CS /Patho/19/012/LPC

दिनांक: 28.02.2019

**कोटेशन सूचना**  
**QUOTATION NOTICE**

Inviting Quotations for Purchase of **Four Channel Coagulation Analyzer** for Department of Pathology & Lab Medicine, AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Four Channel Coagulation Analyzer for** Department of Pathology & Lab Medicine, AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer, 2<sup>nd</sup> floor, Medical College Building Gate No. 05, Tatibandh, G.E. Road, AIIMS, Raipur** up to 06/03/2019 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

क्र. सं. S. No.	सामग्री का विवरण Description of Items	एचएसएन कोड HSN Code	मेक एवं ब्रांड Make / Brand	मात्रा Qty	इकाई दर रु. में Unit Rate in Rs.	टेक्स GST@ %	कर के साथ दर Unit Rate with TAX	कुल मूल्य Total Price
1	<b><u>Four Channel Coagulation Analyzer</u></b>			01				

**Technical Specification for Four Channel Coagulation Analyzer:**

1. Should be an opto-mechanical coagulation analyser which applies the turbo-densitometric measuring principle. All routine clotting tests such as PT, APTT and Fibrinogen assays.
2. The instrument should be operated according to the opto-mechanical measuring principle. Especially suited for lipemic and/or icteric coloured samples.
3. The instrument should be able to perform at least 4 measurements in one run.
4. The instrument should be able to incubate and mix sample and reagent automatically and indicate when to add starting reagent. The auto-start function should automatically detect when to add reagent and trigger mixing and clot detection.
5. The instrument should be based on turbo-densitometry. The lamp intensity should automatically adjust to sample turbidity. Results should be calculated automatically for PT and Fibrinogen.
6. There should be an internal printer, there should be provision of connection to external printer for printing A4 size reports.

7. Ambient temperature for the instrument should be around 10 – 30°C.
8. Relative humidity should be up to 85%.
9. Temperature control Detector and incubator should be around 37.4°C ± 0.3°C.
10. The instrument should have at least 4 wells of detector.
11. The instrument should have reagent holder as; at least 4 positions for 37°C & 1 position for room temperature.
12. The instrument should have at least 16 pre-warmed incubation positions.
13. The instrument should have LCD display.
14. Should have European CE & US FDA certification or BIS approved.
15. The calibration, IQ, OQ and PQ of the instrument should be performed at the time of installation and certificates should be provided.
16. UPS backup adequate for the duration of one cycle of processing should be provided.
17. Start-up kit including cuvettes & reagents for at least 200 tests should be provided free of cost.
18. Support for induction and follow up training of technical staff, on-site standardization and troubleshooting of procedures/ tests to be provided by the company.
19. Two (02) years warranty should be provided.

#### **OTHER COMMERCIAL TERMS**

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. “Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
9. the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
11. The GST registration details may please be furnished.

12. 100% payment against receipt and acceptance of material.
13. Validity of offer should not be less than 90 days
14. No Part supply or Part Payment will be entertained.
15. RTGS detail required for payment purpose.
16. Expenditure will be debitable to GIA-48.
17. Brand & Make should be clearly mentioned in offer (If require).
18. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
19. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

भंडार अधिकारी  
अखिल भारतीय आयुर्विज्ञान संस्थान,  
रायपुर (छ.ग.)