



क्रमांक. AIIMS/R/CS/ENT/631/19/LPC

दिनांक: 27.03.2019

विषय : Inviting Quotations for purchase of Surgical Diathermy and Suction machine for Department of ENT & HNS at AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST No./relevant documents for supply of Surgical Diathermy and Suction machine for Department of ENT & HNS at AIIMS Raipur and should be submitted to office of Stores Officer **2nd Floor, Medical College Building, Gate no. 05** up to 3:00 pm on 08.04.2019 Quotation will be opened same day on 03:30 pm. item description as per detailed below:

S.n. क्र. स.	Name of the Item सामग्री का नाम	Qty मात्रा	Make/ Model	HSN code एच.एस.एन कोड	UNIT RATE IN Rs. इकाई दर रु में	GST@% जी.एस.टी@%	Unit rate incl. GST यूनिट दर जी.एस.टी सहित	Total amount in Rs कुल मूल्य
01	Suction Machine (Specification as per Annexure-I)	01						
02	Surgical Diathermy - Alpha 300 (Specification as per Annexure-I)	01						
Total Value:								

नियम व शर्तें:

1. Rate should be mentioned in words & figure both.
2. L-1 will be decided on total value only.
3. Taxes, if any (should be clearly mention).
4. Delivery Schedule – within 15 days from the date of issue of PO.
5. Price should be FOR Destination basis.(i.e. concerned department)
6. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
7. Quotation No/Name and Due date of opening must be written on top of envelop.
8. GST rates applicable on your quoted item may please be confirmed. **HSN code** for each item should be clearly mention.
9. Please confirm if there any change (Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. “Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. The GST registration details may please be furnished.
11. 100% payment against receipt and acceptance of material.

12. Validity of offer should not be less than 90 days.
13. No Part supply or Part Payment will be entertained.
14. RTGS detail required for payment purpose.
15. Expenditure will be debitable to GIA-GENERAL.
16. Brand & Make should be clearly mentioned in offer (If require).
17. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.
18. The Quantity of above column is totally tentative. It may be increased or decreased at the time of placement of order.
19. **Technical Compliance report must be submitted**

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)

Technical specification

Suction Machine

Salient Features :-

- Maintenance free oil less, piston cylinder pump with high flow without pollution by oil mist.
- Ergonomically designed ABS body cover.
- Compact size wide mouth glass bottles with stopper and silicon seal ring, convenient to open and clean.
- Effective overflow protection device to prevent liquids entering into the pump.
- Hoses and connections made of transparent, non-toxic material are transparent and easy to observe and link.
- Parallel hand and foot control can be selected of choice.
- Unique bacterial filter to prevent air borne contamination in sterile rooms.
- Reliable and safe operation.
- Ergonomically designed ABS body cover.

Specification :-

- Maximum Vacuum : ≥ 0.09 Mpa (680 mmHg)
- Adjustable Range : 150 mmHg ~Max. Vacuum
- Flow Rate : ≥ 40 LPM
- Noise : ≤ 60 dB(A)
- Storage Bottle : 250ml \times 2 (Glass jars)
- Power supply : 220V, $\pm 10\%$, 50 Hz
- Input : 250
- Weight : GW -22.5 Kg, NW - 20 Kg
- Dimensions : 465 \times 425 \times 920mm

Surgical Diathermy - Alpha 300

Technical Data

Feature

Alpha 300

Monopolar Cut

Pure Cut	Watt
Blend 1	Watt
Blend 2	Watt

Power in 300 Ohms

300
200
100

Monopolar Coagulation

Desiccat	Watt
Fulgurate	Watt
Spary	Watt

Power in 500 Ohms

100
100
100

Bipolar Coagulation

Precise	Watt
Micro	Watt

Power in 125 Ohms

70
40

Power in 300 Ohms

Bipolar Cut	Watt
-------------	------

120

N.B. Output power tolerance +/-15% or 5W whichever is greater.

No. of Programs	-
Power fail memory	-
Power Supply	VAC 50Hz
External power supply	-
Power Consumption	VA Max
Dimensions (WxHxD)	mm
Weight Approx	Kg

10
Yes
230
SMPS
750
380×133×390
6.2

Standard Accessories

Disposable Hand Switch	2No.	Yes
Patient Plate	2Silicon Split	Yes
Bipolar forcep	2	Yes
Foot Switch	2 Pedal + 1 Pedal	Yes

- Bipolar bayonet forceps 05 No.
- Straight forceps 05 No.