



AIIMS/R/CS /Micro/18/79/A/LPC

Date:23.05.2019

विषय/Sub:Inviting Quotations for procurement of Micropipette Single Channel for Microbiology Department, at AIIMS Raipur.

कोटेशन सूचना QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having GST Number and relevant documents for Micropipette Single Channel for Microbiology Department,at AIIMS Raipur. The quotation should be submitted to Medical College Building,2nd Floor,Gate no.05 office of Store Officer up to 3:00pm on 29.05.2019. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

क्र.स. S.No.	सामग्री का विवरण (Description of Items)	आवश्यक मात्रा Requirement	ब्रांड Make/ Model	HSN Code	इकाई दर ₹ में UNIT Rate in Rs.	GST	कुल रकम Total Amount With GST in Rs.
1	Micropipette Single Channel Volum: 0.5/1 to 10 µl	02 Nos					
2	Micropipette Single Channel Volum: 10 to 100 µl	02 Nos					
3	Micropipette Single Channel Volum: 100 to 1000 µl	02 Nos					

Specifications for Adjustable Single Channel Micropipette.

- 1.Micropipette should be adjustable with ultra- light weight.
- 2.Micropipette should be fully autoclavable.
- 3.Adjustment opening for adjusting pipettes to a specific liquid and volume.
- 4.Piston System: Ultra-light system and should be made of fortom.
- 5.It should be highly resistant to heat,acids and alkali.mildew,bleaches,aging,sunlight and abrasion.
- 6.No discoloration upon UV irradiation.
- 7.It should have button to control very low operating force,Color indication for pipette volume,Positioned for perfect Ergonomics.
- 8.Volume Display:Four digit display with magnifier and 2 Button operations.
- 9.Spring loaded tip cone enable to improved ergonomics to reduce stress without sacrificing tightness.
- 10.Ejector:Very low ejection force and positioned for perfect ergonomics.
- 11.Quick connection clip:Remove lower part easily.
- 12.Viable calibration seal to indicate factory calibration not changed.
- 13.Vendor should have NABL calibration facility,The proof for the same has to be enclosed along with the Tender.
- 14.Each micropipette should be supplied with factory calibration certificate.
- 15.Each Micropipette upto 1 ml volume should have autoclavable tip box with 96 tips.
- 16.Micropipette should be compitable with tips of any other manufacturer.
- 17.The item should be ISO 8655 and IVD compliant.

Technical Data:

a. 0.5/1-10 µ L adjustable micropipette.

Qty:02 Nos

Volume range – 0.5/1 – 10 µL		
Volume	Systematic error (approx.)	Random error (apprx.)
0.5 µL	± 0.04 µL	± 0.025 µL
1 µL	± 0.025 µL	± 0.02 µL
5 µL	± 0.075 µL	± 0.04 µL

10 µL	± 0.1 µL	± 0.05 µL
-------	----------	-----------

b. 10 – 100 µ L adjustable micropipette.

Qty:02 Nos

Volume range – 10 – 100 µL		
Volume	Systematic error (approx.)	Random error (apprx.)
10 µL	± 0.03 µL	± 0.1 µL
50 µL	± 0.5 µL	± 0.15 µL
100 µL	± 0.8 µL	± 0.2 µL

c. 0.5/1-10 µ L adjustable micropipette.

Qty:02 Nos

Volume range – 100 – 1000 µL		
Volume	Systematic error (approx.)	Random error (apprx.)
100 µL	± 3.0 µL	± 0.6 µL
500 µL	± 5.0 µL	± 1 µL
1000 µL	± 6.0 µL	± 2 µL

Note:1.Quotation without technical catalogue will not be considered.

2.Items should be available for demonstration.

3.Should come with three years warranty from the date of supply.

4.Supplier should ensure the service and supply of spares beyond the warranty period.

नियम एवं शर्तें

Terms & Conditions.

1. Rate should be mentioned in words & figures both.
2. **GST rates applicable** on your quoted item may please be confirmed.
3. Delivery Schedule- within 15 days from the date of issue of P.O.
4. No additional documents related to this NIQ will be entertained after opening of NIQ.
5. Price should be F.O.R. for Destination basis (i.e.Concerned Department)
6. LD@0.5% of delayed supply per week or part week for delay of supply of material subject to maximum upto 10% of delayed supply should be deducted.
7. Quotation No/Name and Due date of opening must be written on top of the envelop.
8. **Brand & Make & warranty should be clearly mentioned in offers as well as tender/quotation specific authorization may be submit with the offer/bid.**
9. **The GST registration details may please be furnished..**
10. Please confirm if there any change (Upward/Reduction) in your **Basic Price** structure. And you are also requested to pass the Input Credit as per the following **Anti Profitering Clause** of GST.“**Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices**”
11. RTGS details required for payment purpose.
12. 100% payment against receipt and acceptance of material.
13. No part supply or part payment will be entertained.
14. Validity of offer should not be less than 90 days.
15. Supply,Installation and Commissioning will be done by firm (if applicable).
16. The quantity shown in above column are totally tentative, it can be increase and decrease at the time of placement of purchase order.
17. AIIMS Raipur reserved the rights to place order for full or part quantity to one or more firm.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)

Vendor Details

Name	
Aadhaar No. (if any)	
PAN	
GST.NO.	
Address	
City	
State	
Pincode	
Mobile No.	
Phone No.	
E-mail	
Bank Name	
Bank A/c No.	