

<p>a. It should be radiopaque tis attached to clear bodies with separate vent lines.</p> <p>b. These cannula should be able to monitor aortic root pressure and left heart venting.</p> <p>c. It should be supplied with a stainless steel introducer.</p>									
5. Cannula Aortic root with vent line 12 Ga	5	Nos							
6. Cannula Aortic root with vent line 14 Ga	5	Nos							
7. Cannula aortic root with vent line 16 Ga	4	Nos							
<p>Cannula DLP malleable single stage venous cannula</p> <p>a. It should have multiport tips and malleable ,one piece ,kink resistant wirewound bodies with depth markings.</p> <p>b. An integrated malleable wire outside the wire winding optimize the inner diameter for maximum performance.</p>									
8. Cannula SS venous 26 Fr	4	Nos							
9. Cannula SS venous 28 Fr	5	Nos							
10. Cannula SS venous 30 Fr	5	Nos							
11. Cannula SS venous 32 Fr	5	Nos							
12. Cannula SS venous 34 Fr	5	Nos							
13. Cannula SS venous 36 Fr	4	Nos							
<p>DLP single stage venous cannula with right angle metal tip</p> <p>a. Cannula should have beveled metal ,multiport tips</p> <p>b. Kink resistant wirewound bodies with a tip orientation line.</p>									
14. Cannula Ven Rt metal 20 Fr	5	Nos							
15. Cannula Ven Rt metal 22 Fr	5	Nos							

16. Cannula Ven Rt metal 24 Fr	5	Nos								
17. Cannula Ven Rt metal 28 Fr	5	Nos								
DLP Metal tip Arterial cannula a. It should have beveled metal tips and tapered bodies with tip orientation line with soft flange										
18. Cannula ART metal tip 20 Fr	5	Nos								
19. Cannula ART metal tip 22 Fr	5	Nos								
20. Cannula ART metal tip 24 Fr	5	Nos								
Two stage Venous cannula A. it should have multiport tips with atrial baskets and kink resistant. B. it should have wirewound bodies with depth marking.										
21. Cannula two stage Ven 36/46 Fr	5	Nos								
22. Cannula two stage Ven 32/40 Fr	3	Nos								
23. Cannula two stage Ven 36/51 Fr	5	Nos								
Grand Total										

Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Products are certified from **BIS/CE/USFDA & ISO** as applicable, the Certificate to this effect should be attached.
5. Supply should be done within 15 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from Representative of MS.
8. **Quotation Name/No. and due date of opening must be mentioned on top of envelopes.**

9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
11. All other terms & condition as per GFR 2017.
12. Material to be delivered at **CTVS Department, AIIMS Raipur.**
13. **Validity of the quotation should be 90 days from the date of opening.**
14. **Firm to submit documentary evidence in support of their claim for GST at the time of submission of bills.**
15. **Firm to submit sample if required.**

Stores Officer (H)
AIIMS, Raipur
(C.G)