



क्रमांक. AIIMS/R/CS/Neuro/483/19/LPC

दिनांक: 13.07.2019

विषय : Inviting Quotations for purchase of Patient Connective Warming Device for Neurosurgery Department at AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST No./relevant documents for supply of of Patient Connective Warming Device for Neurosurgery Department at AIIMS Raipur and should be submitted to office of Stores Officer **2nd Floor, Medical College Building, Gate no. 05** up to 3:00 pm on 19.07.2019 item description as per detailed bellow:

S.n. क्र. स.	Name of the Item सामग्री का नाम	Qty मात्रा	Make / Model	HSN code एच.एस.एन कोड	UNIT RATE IN Rs. इकाई दर रु में	GST@% जी.एस.टी @%	Unit rate incl. GST यूनिट दर जी.एस.टी सहित	Total amount in Rs कुल मूल्य
01	Patient Connective Warming Device (Specification as per Annexure-I)	02 No.						
Total Value:								

नियम व शर्तें:

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 15 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. GST rates applicable on your quoted item may please be confirmed. **HSN code** for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. “Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
9. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
10. The GST registration details may please be furnished.
11. 100% payment against receipt and acceptance of material.
12. Validity of offer should not be less than 90 days
13. No Part supply or Part Payment will be entertained.
14. RTGS detail required for payment purpose.
15. Expenditure will be debitable to GIA-GENERAL.
16. Brand & Make should be clearly mentioned in offer (If require).
17. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.
18. The Quantity of above column is totally tentative. It may be increased or decreased at the time of placement of order.
19. **Technical Compliance report must be submitted.**

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)

Annexure I

Specification of Patient Connective Warming Device

- Should be suitable for intra operative & Pre-post-operative application.
- Should have set temperature.
- A. High: $43 \pm 1.5^{\circ}\text{C}$ ($109.4 \pm 2.7^{\circ}\text{F}$)
- B. Med: $38 \pm 1.5^{\circ}\text{C}$ ($100.4 \pm 2.7^{\circ}\text{F}$)
- C. Low $32 \pm 1.5^{\circ}\text{C}$ ($89.6 \pm 2.7^{\circ}\text{F}$)/Room Temperature
- Hose end temperature sensing increased airflow
- Microprocessor control: 3 temperature sensor
- Built-in Hours meter makes to monitor usage for preventive maintenance (desirable)
- Temperature setting to track and report forced air warming therapy duration.
- Over temperature, calibration and fault codes servicing right on the front panel, no need to open the unit.
- Should have safety alarms like over temperature.
- Compact and light weight.
- Control unit weight should not be more than 5.5 kg.
- Dimension: 22 cm length x 22 cm width x 28 cm Height.
- Air filter: high efficacy 0.2 micron.
- Self-life of filter should be at least 1000 hours.
- USFDA Approved.
- 02 Year warranty