अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़



(All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

No. AIIMS/R/CS/39505/OPTH/19/LPC

Date: 04/10/2019

Notice for inviting Quotation for supply of <u>Non-Consumable items i.e. Lensmeter for Department</u> <u>of Ophthalmology, AIIMS, Raipur.</u>

<u>कोटेशन सूचना</u> <u>QUOTATION NOTICE</u>

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST No./relevant documents for **supply of Lensmeter for Department of Ophthalmology** to be used in AIIMS, Raipur and should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on **11.10.2019**, item description as per detailed bellow:

S.no.	Item Detail	Qty.	Make	Unit	HSN	Unit	GST	Total
		Req.			Code	Price	@	Amount
1.	Lensmeter							
	Specification:	01						
	1. Certification: BIS / ISO / European CE / US FDA							
	approved product.							
	2. Measurement Range:							
	I. Sphere : 0 to ± 25D (Steps of 0.01/0.12/0.25 D)							
	II. Cylinder : 0 to ± 10 D (Steps of 0.01/0.12/							
	0.25D)							
	III. Axis : 1 0 to 180 0 (Steps of 1 0)							
	IV. Addition : 0 to ±10.0D (Steps of 0.01/0.12/							
	0.25D)							
	V. Prism : 0 to 10Δ (Horizontal, Vertical)							
	3. Cylinder mode : Mix -/+							
	4. Prism Mode : X – Y (Orthogonal Coordinates), P							
	-B (Polar Coordinates),no display							
	5. UV Transmittance measurement : $0 \sim 100\%$							
	(1% / 5% steps)							
	6. Lens Detection : Automated for single & amp;							
	Progressive lenses 7. Contact lens Measurement : Hard / Soft contact							
	lens							
	8. Light source :							
	I. Refractive Measurement : Green LED 535 ±5 nm							
	II. UV Measurement : UV LED $365 \pm 5 \text{ nm}$							
	9. Wavelength Criterion : e-line/d-line with ABBE							
	compensation							
	10. Lens Diameter : 5 to 100 mm							
	11. PD measurement range : upto 90 mm							
	12. Display System:							
	I. Color touch screen panel							
	II. Screen size: minimum 5.5inches							
	III. Wider angle of observation							
	13. External Interface : Rs-232 C							
	14. Power supply : 100-240 V / 50 /60 Hz 1.2A							
	(AC Adapter) with having Auto							
	power save mode							
	15. Dimensions : 170~200 (W) x 200~225 (D) x							
	400~425 (H) mm							
	16. Weight : 3.5 kg ~ 5 kg							
	17. Printer should be included							
	18. Motorised table should be included							
	Warranty -05 year							

नियम और शर्ते:

Terms and Condition:

- 1. Rate should be mentioned in word and figures both.
- 2. Delivery Schedule within 15 days from the date of issue of PO.
- 3. Price should be F.O.R. Destination basis (Department of Ophthalmology, AIIMS Raipur).
- 4. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
- 5. Quotation No./Name and Due date of opening must be written on top of envelop.
- 6. GST rates applicable on your quoted item may please be confirmed. HSN code should be clearly mentioned.
- 7. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.

"Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices".

- 8. The GST registration details may please be furnished.
- 9. RTGS details required for payment purpose.
- 10. 100% payment against received and acceptance of material.
- 11. No part supply or part payment will be entertained.
- 12. No claim will be entertained regarding interest on any payment of any circumstance.
- 13. Supply, installation and commissioning will be done by firm (if applicable).
- 14. Where there is statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the supplier at rates as notified from time to time.
- 15. No payment shall be made for rejected stores. Rejected equipment's must be removed by the supplier within two weeks of the dated of issue of rejection advice at their own cost & replace immediately. In case these are not removed these will auctioned at the risk and responsibility of the suppliers without notice.
- 16. Brand and Make should be clearly mentioned in offer as well as tender/quotation specific authorization may be submit with the offer/Bid.
- 17. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
- 18. Validity of offer should not be less than 90 days.
- 19. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.

Sr. Administrative Officer AIIMS, Raipur