



AIIMS/R/CS /Patho/19/170/LPC/

Date: 29.11.2019

QUOTATION NOTICE

Inviting Quotations for Purchase of **Hot Air Oven** for use in emergency lab in Department of Pathology & Lab Medicine, AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Hot Air Oven** for use in emergency lab in Department of Pathology & Lab Medicine,, AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer, 2nd floor, Medical College Building Gate No. 05, Tatibandh, G.E. Road, AIIMS, Raipur** up to 09/12/2019 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

क्र. सं. S.No.	सामग्री का विवरण Description of Items	एचएसए न कोड HSN Code	मेक एवं ब्रांड Make / Brand	मात्रा Qty	इकाई दर रु. में Unit Rate in Rs.	टेक्स GST	कर के साथ दर Unit Rate with TAX	कुल मूल्य Total Price
1	Hot Air Oven Technical Specification as per given below-			01 Set				

Technical Specification for Hot Air Oven:

- It should have Chamber capacity of 250 ltr or more.
- It should have internal dimensions (L x D x H): 600 x 600 x 700 mm approx.
- It should have Exterior construction of SS 304.
- Chamber should be at least 1 mm thickness.
- Oven Interior construction should be of stainless steel 304 grade.
- It should have door construction having double tempered sandwich glass window to enable user for easy viewing of samples without door opening.
- It should be provided with special Incoloy sheathed spread out shaped heater for uniformity and efficient heating.
- Oven should be provided with Impellor type Aluminium blower for internal circulation for better uniformity.
- It should be provided with non-contact type door switch for auto cut off when door opens during operation.
- It should have fitted with stainless steel vertical channels (CNC Press Punched) for height adjustment of trays in steps of 25mm.
- It should have seamless round cornered internal chamber ensures easy cleaning.
- It should have solid and plain bottom without electrical.
- It should have Digital PID temperature controller with PT100 sensor for precise monitoring & control

- Controller should be Equipped with timer (timer bypass option should be available), auto tuning and alarms.
- Oven Temperature range should be ambient +5°C to 250°C
- Control accuracy should be $\pm 0.5^\circ\text{C}$
- It should have independent over temperature safety protection and should be user settable.
- It should have over current protection.
- Oven should be CE certified.
- Supply Voltage should be 230 V AC, 50 HZ, Single phase.
- 5 KVA stabilizer should be provided.
- Oven should be CE certified or BIS approved
- The calibration, IQ, OQ and PQ of the instrument should be performed at the time of installation and certificates should be provided.
- Support for induction and follow up training of technical staff, on-site standardization and troubleshooting of procedures/ tests to be provided by the company.
- Three (03) years warranty should be provided.

OTHER COMMERCIAL TERMS

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
9. The benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. The GST registration details may please be furnished.
11. 100% payment against receipt and acceptance of material.
12. Validity of offer should not be less than 90 days
13. No Part supply or Part Payment will be entertained.
14. RTGS detail required for payment purpose.
15. Expenditure will be debitable to GIA-48.
16. Brand & Make should be clearly mentioned in offer (If require).
17. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
18. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

**Sr. Administrative Officer
AIIMS, Raipur (C.G.)**