



AIIMS/R/CS /Patho/19/198/LPC/

Date: 09.12.2019

QUOTATION NOTICE

Inviting Quotations for Purchase of **Homogenizer** for use in molecular pathology lab in Department of Pathology & Lab Medicine, AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Homogenizer** for use in molecular pathology lab in Department of Pathology & Lab Medicine,, AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer, 2nd floor, Medical College Building Gate No. 05, Tatibandh, G.E. Road, AIIMS, Raipur** up to 18/12/2019 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

क्र. सं. S.No.	सामग्री का विवरण Description of Items	एचएसए न कोड HSN Code	मेक एवं ब्रांड Make / Brand	मात्रा Qty	इकाई दर रु. में Unit Rate in Rs.	टेक्स GST	कर के साथ दर Unit Rate with TAX	कुल मुल्य Total Price
1	Homogenizer Technical Specification as per given below-			01 Set				

Technical Specification for Homogenizer:

- Instrument should be ideal homogenizer for extraction from samples like bacterial/ yeast cells, plant tissues etc. for molecular analysis.
- Instrument can accommodates wide range of screw cap homogenizing tubes sizing 3 x 2 or more.
- Instrument should have powerful maintenance free motor- Homogenizes most challenging samples.
- Instrument should be designed to generate less heat and convenient loading.
- Instrument should have variable cyclic intervals for effective homogenization.
- Instrument should features an integrated lid lock for user safety by preventing lid opening while operation.
- Instrument should have easy user interface with large digital display.
- Speed range should be 2900 rpm to 3900 rpm or better.
- Run time should be at least 3 seconds to 160 seconds.
- Pulse timer should be at least 3 seconds to 99 seconds.
- Power supply for same should be 230 VAC/50Hz.
- The system should comply with European CE, US FDA or BIS approved.
- The calibration, IQ, OQ and PQ of the instrument should be performed at the time of installation and certificates should be provided.

- Document supporting track record and satisfactory performed from institutes of national importance (minimum one) should be provided.
- Support for induction and follow up training of technical staff, on-site standardization and troubleshooting of procedures/ tests to be provided by the company.
- 01-year warrantee should be provided.

OTHER COMMERCIAL TERMS

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
9. The benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. The GST registration details may please be furnished.
11. 100% payment against receipt and acceptance of material.
12. Validity of offer should not be less than 90 days
13. No Part supply or Part Payment will be entertained.
14. RTGS detail required for payment purpose.
15. Expenditure will be debitable to GIA-48.
16. Brand & Make should be clearly mentioned in offer (If require).
17. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
18. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

**Sr. Administrative Officer
AIIMS, Raipur (C.G.)**