



No.: AIIMS/R/CS/632/BPS/19/LPC/

Date: 14-12 -2019

QUOTATION NOTICE

Inviting quotations for purchase of “**Mobile Examination Light**” for Department of Burn & Plastic Surgery at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2nd Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **20-12-2019** before **3:00 pm**. The quotations will be opened on the same day at **3.30 pm**.

Terms & Condition:

1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed "**Technical Bid for Mobile Examination Light**" and "**Financial Bid for Mobile Examination Light**". Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as "**Quotation for supply of Mobile Examination Light**".
3. **All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.**
4. Rate should be mentioned in words & figure both.
5. Taxes/GST, if any (Kindly mention in table)
6. Delivery schedule: within 15 days from the date of issue of P.O.
7. No additional documents related to this NIQ will be entertained after opening of NIQ.
8. Price should be FOR Destination basis (i.e. Department of Burn & plastic Surgery, AIIMS Raipur).
9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
10. Quotation No/Name and due date of opening must be written on top of envelop.
11. **Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.**
12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned.**
13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.
“**Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices**”.
14. The GST registration details may be furnished.
15. RTGS details required for payment purpose.
16. 100% payment against receipt and acceptance of material.
17. No part supply or part payment will be entertained.
18. Validity of offer should not be less than 90 days.
19. Supply, installation and commissioning will be done by firm (if applicable).
20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.

**Sr. Administrative Officer,
AIIMS, Raipur (C.G.)**



ANNEXURE-I

Ref. No. & Date: -
 Quotation Notice No.: -
 Brand/Make Name:-
 Model No.:-
 Warranty-

TECHNICAL BID

Purchase of “**Mobile Examination Light**” for Department of Burn & Plastic Surgery at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	Offered Specification	Specify Technically Complied / Not Complied/ if any Deviations
1.	Mobile Examination Light: 1. The Central illumination should be at least 60,000 lux at 1Mtr distance 2. Dimming provision should be between 30-100% 3. Light field diameter should be at least 170 mm 4. The color rendering index Ra and Red rendering index R9 should be > 95 5. The light should be based on reflector technology for better shadow resolution 6. The lifetime of LEDs should be minimum 50,000 hrs 7. The color temperature should be 4500 K 8. Should have scratch proof safety glass 9. Illumination depth at least 900 mm 10. Light wt. aluminum housing for better cool light 11. Light head wt should <= to 3 kg 12. Total illumination strength (W/m ²) should 120 13. FDA Listing or Approval and European CE /BIS Declaration to be enclosed along with the bid. 14. Protection class should be IEC 60601-1 certified to be enclosed along with the bid 15. Rated Power (W) : 15W 16. It should have minimum of 18 LEDs to create a perfect shadow-less light. 17. The examination light should be mounted on castors for easy transportation. Warranty -01 years.	01 Nos.		

Sign of bidder :-
 Date :-
 Name of the bidder :-
 Seal of the bidder :-



ANNEXURE-II

Ref. No. & Date: -

Quotation Notice No.: -

FINANCIAL BID

Purchase of “**Mobile Examination Light**” for Department of Burn & Plastic Surgery at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	HSN Code	Make/ Brand/ Model	Unit Price in ₹.	GST Per Unit @	Total Unit Price with Tax in ₹.	Total Amount in ₹.
1.	Mobile Examination Light: 1. The Central illumination should be at least 60,000 lux at 1Mtr distance 2. Dimming provision should be between 30-100% 3. Light field diameter should be at least 170 mm 4. The color rendering index Ra and Red rendering index R9 should be > 95 5. The light should be based on reflector technology for better shadow resolution 6. The lifetime of LEDs should be minimum 50,000 hrs 7. The color temperature should be 4500 K 8. Should have scratch proof safety glass 9. Illumination depth at least 900 mm 10. Light wt. aluminum housing for better cool light 11. Light head wt should <= to 3 kg 12. Total illumination strength (W/m ²) should 120 13. FDA Listing or Approval and European CE /BIS Declaration to be enclosed along with the bid. 14. Protection class should be IEC 60601-1 certified to be enclosed along with the bid 15. Rated Power (W) : 15W 16. It should have minimum of 18 LEDs to create a perfect shadow-less light. 17. The examination light should be mounted on castors for easy transportation. 18. Warranty -01 years.	01 Nos.						

Sign of bidder :-

Date :-

Name of the bidder :-

Seal of the bidder :-