

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, GE Road, Raipur-492 099 (CG)

<u>www.aiimsraipur.edu.in</u>

No.: AIIMS/R/CS/Pul_Med/Capital/19-20/101/

Date: 29/01/2020

Inviting quotations for purchase of **"DVT Pump"** for Department of Pulmonary Medicine at AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2^{nd} Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **03/02/2020** before **3:00** pm. The quotations will be opened on the same day at **3.30 pm**.

Terms & Condition:

- 1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
- The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed "Technical Bid for DVT Pump" and "Financial Bid for DVT Pump". Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as "Quotation for supply of DVT Pump".
- 3. All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.
- 4. Rate should be mentioned in words & figure both.
- 5. Taxes/GST, if any (Kindly mention in table)
- 6. Delivery schedule: within 15 days from the date of issue of P.O.
- 7. No additional documents related to this NIQ will be entertained after opening of NIQ.
- 8. Price should be FOR Destination basis (i.e. Department of Pulmonary Medicine, AIIMS Raipur).
- 9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
- 10. Quotation No/Name and due date of opening must be written on top of envelop.
- 11. Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.
- 12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned**.
- 13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.

"Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices".

- 14. The GST registration details may be furnished.
- 15. RTGS details required for payment purpose.
- 16. 100% payment against receipt and acceptance of material.
- 17. No part supply or part payment will be entertained.
- 18. Validity of offer should not be less than 90 days.
- 19. Supply, installation and commissioning will be done by firm (if applicable).
- 20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.

Sr. Administrative Officer, AIIMS, Raipur (C.G.)



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

ANNEXURE-I

Ref. No. & Date: -Quotation Notice No.: -Brand/Make Name:-Model No.:-Warranty-

TECHNICAL BID

Purchase of **"DVT Pump"** for Department of Pulmonary Medicine at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	Offered Specification	Specify Technically Complied / Not Complied/ if any
				Deviations
N.	 DVT Pump: It should have LCD Display of all Parameters time, pressure, battery life etc. It should have Auto Optimum Operation by one touch button. It should have selectable manual settings operation system (Pressure interval time). It should have safety alarm system by self- detection system (cuff, pressure and power battery). It should have stable compression cycle operation by regular vascular cycle time applicable. It should have self pressure sensor calibration during power input and operation intermittently. It should have silent operation noise minimized for comfortable used in operation or ICU room. It should have Power Consumption ; 35 W. It should have pressure rang (mmHg): Thigh/Bloot/calf-20,30,40,50,60, mmHg, Foot 120,130,140 mmHg. It should have main body : 250 (W) x 140 (D) x 245 (H)mm 			Not Complied/ if any
	 140 (D) x 245 (H)mm 13. It should have weight : 3.5 Kg (battery included) 14. It should have battery life time 6-8 hrs. 15. It should have be BIS/ISO/European CE/ USFDA approved. 16. Warranty: 01 Year. 			

Sign of bidder:-Date:-Name of the bidder:-Seal of the bidder:-



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

ANNEXURE-II

Ref. No. & Date: -Quotation Notice No.: -

FINANCIAL BID

Purchase of **"DVT Pump"** for Department of Pulmonary Medicine at AIIMS Raipur. Details of item are given as under:

S.	Description of item(s)	Qty./	HSN	Make/	Unit	GST Per	Total Unit	Total
N.	Description of rem(6)	Unit	Code			Unit @	Price with	Amount in
				Model	₹.		Tax in ₹.	₹.
1.	DVT Pump:							
	1. It should have LCD Display							
	of all Parameters time,							
	pressure, battery life etc.							
	2. It should have Auto Optimum							
	Operation by one touch button.							
	3. It should have selectable manual							
	settings operation system							
	(Pressure interval time).							
	4. It should have safety alarm							
	system by self- detection system							
	(cuff, pressure and power battery).							
	5. It should have stable compression							
	cycle operation by regular							
	vascular cycle time applicable.							
	6. It should have self pressure sensor							
	calibration during power input	02						
	and operation intermittently. 7. It should have silent operation	02 Nos.						
	noise minimized for comfortable	1105.						
	used in operation or ICU room.							
	8. It should have voltage: AC 100-							
	240v, 50/60Hz.							
	9. It should have Power							
	Consumption ; 35 W.							
	10. It should have pressure rang							
	(mmHg): Thigh/Bloot/calf-							
	20,30,40,50,60, mmHg, Foot							
	120,130,140 mmHg.							
	11. It should have interval time:							
	24,48,60 second.							
	12. It should have main body : 250							
	(W) x 140 (D) x 245 (H)mm							
	13. It should have weight : 3.5 Kg							
	(battery included)							
	14. It should have battery life time 6-							
	8 hrs.							
	15. It should have be							
	BIS/ISO/European CE/ USFDA							
	approved.							
	16. Warranty: 01 Year.							

Sign of bidder:-Date:-Name of the bidder:-Seal of the bidder:-