



No.: AIIMS/R/CS/Pul_Med/Capital/19-20/102/

Date:03/02/2020

QUOTATION NOTICE

Inviting quotations for purchase of “**Portable Machine for Testing Exercise Capacity**” for Department of Pulmonary Medicine at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2nd Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **10/02/2020** before **3:00** pm. The quotations will be opened on the same day at **3.30 pm**.

Terms & Condition:

1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed “**Portable Machine for Testing Exercise Capacity**” and “**Financial Bid for “Portable Machine for Testing Exercise Capacity”**”. Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as **Quotation for supply of “Portable Machine for Testing Exercise Capacity”**.
3. **All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.**
4. Rate should be mentioned in words & figure both.
5. Taxes/GST, if any (Kindly mention in table)
6. Delivery schedule: within 15 days from the date of issue of P.O.
7. No additional documents related to this NIQ will be entertained after opening of NIQ.
8. Price should be FOR Destination basis (i.e. Department of Pulmonary Medicine, AIIMS Raipur).
9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
10. Quotation No/Name and due date of opening must be written on top of envelop.
11. **Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.**
12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned.**
13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
14. The GST registration details may be furnished.
15. RTGS details required for payment purpose.
16. 100% payment against receipt and acceptance of material.
17. No part supply or part payment will be entertained.
18. Validity of offer should not be less than 90 days.
19. Supply, installation and commissioning will be done by firm (if applicable).
20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.

**Sr. Administrative Officer,
AIIMS, Raipur (C.G.)**



ANNEXURE-I

Ref. No. & Date: -
Quotation Notice No.: -
Brand/Make Name:-
Model No.:-
Warranty-

TECHNICAL BID

Purchase of “**Portable Machine for Testing Exercise Capacity**” for Department of Pulmonary Medicine at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	Offered Specification	Specify Technically Complied / Not Complied/ if any Deviations
1.	Portable System for Exercise Testing Capacity- 1. The system should be light weight with in-built display and meets American Thoracic Society recommendation. 2. The System should have facility for real – time recording and display of 6- minute walk distance, Heart rate, and SpO2 recording during walking 3. Options for multiple type of oximetry sensors. 4. It should have internal memory of at least 100 tests. 5. It should allow manual entry of pre-test and post-test parameters (Dyspnea & fatigue measured on Borg scale, blood pressure, distance walked, etc.) 6. It should have facility to down load tests to the computer with window 10. 7. It should be US- FDA or European CE approved. Essential Accessories- 1. Two number of each small, medium and large sensor. 2. One number of each ear clip sensor, reflectance sensor.	01 Set		

Sign of bidder :-
Date :-
Name of the bidder :-
Seal of the bidder :-



ANNEXURE-II

Ref. No. & Date: -

Quotation Notice No.: -

FINANCIAL BID

Purchase of “**Portable Machine for Testing Exercise Capacity**” for Department of Pulmonary Medicine at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	HSN Code	Make/ Brand/ Model	Unit Price in ₹.	GST Per Unit @	Total Unit Price with Tax in ₹.	Total Amount in ₹.
1.	Portable System for Exercise Testing Capacity- 1. The system should be light weight with in-built display and meets American Thoracic Society recommendation. 2. The System should have facility for real – time recording and display of 6-minute walk distance, Heart rate, and SpO2 recording during walking 3. Options for multiple type of oximetry sensors. 4. It should have internal memory of at least 100 tests. 5. It should allow manual entry of pre-test and post-test parameters (Dyspnea & fatigue measured on Borg scale, blood pressure, distance walked, etc.) 6. It should have facility to download tests to the computer with window 10. 7. It should be US- FDA or European CE approved. Essential Accessories- 1. Two number of each small, medium and large sensor. 2. One number of each ear clip sensor, reflectance sensor.	01 Set.						

Sign of bidder :-

Date :-

Name of the bidder :-

Seal of the bidder :-