



No.: AIIMS/R/CS/ANAT/366/19/LPC/A

Date: 01.07.2020

QUOTATION NOTICE

Inviting quotations for purchase of “UV Transilluminator” for Anatomy Department at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2nd Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **08.07.2020** before **3:00 pm**.

Terms & Condition:

1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed "**Technical Bid for UV Transilluminator**" and "**Financial Bid for UV Transilluminator**". Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as "**Quotation for supply of UV Transilluminator**" for Anatomy Department.
3. **All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.**
4. Rate should be mentioned in words & figure both.
5. Taxes/GST, if any (Kindly mention in table)
6. Delivery schedule: within 15 days from the date of issue of P.O.
7. No additional documents related to this NIQ will be entertained after opening of NIQ.
8. Price should be FOR Destination basis (i.e. ENT & HNS Department, AIIMS Raipur).
9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
10. Quotation No/Name and due date of opening must be written on top of envelop.
11. **Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.**
12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned.**
13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
14. The GST registration details may be furnished.
15. RTGS details required for payment purpose.
16. 100% payment against receipt and acceptance of material.
17. No part supply or part payment will be entertained.
18. Validity of offer should not be less than 90 days.
19. Supply, installation and commissioning will be done by firm (if applicable).
20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.

**Stores Officer,
AIIMS, Raipur (C.G.)**



ANNEXURE-I

Ref. No. & Date: -
Quotation Notice No.: -
Brand/Make Name:-
Model No.:-
Warranty-

TECHNICAL BID

Purchase of “**UV Transilluminator**” for Anatomy Department at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	Offered Specification	Specify Technically Complied / Not Complied/ if any Deviations
1.	<p>Ultra-violet (UV) transilluminator equipment to be used for visualization of DNAs and proteins agarose and polyacrylamide gels after electrophoresis.</p> <ul style="list-style-type: none">• The system should be extremely protected and covered from all sides.• Should have slot to place the gel for illumination and use hands from both the side to adjust the gel.• System should have provision to use the camera.• System should have metal body cover to limit or avoid direct exposure with the radiation.• The Transmission Size(W*L) should be not less than 200*150mm• The Transmission UV Wavelength of the system should be 302nm• Reflectance UV Wavelength should be 254, 365nm• The Transmission UV Lamp Power should be 302nm(8W)• The Reflectance UV Lamp Power should be 254nm(11W), 365nm(11W)• The system should have operating port for cutting the gel.• System should do viewing without requiring a dark room.• UV filter should be of Glass Quartz ultraviolet lamp.• The lamp should have a minimum self-life of 5 years and uniformity of the light should not vary with time.• The required Power Supply to run the system should be AC110/220V±10%, 50/60Hz• The dimension of the system not be more than (W*D*H) 435*295*490 mm•	01 Nos.		

Sign of bidder :-
Date :-
Name of the bidder :-
Seal of the bidder :-



ANNEXURE-II

Ref. No. & Date: -

Quotation Notice No.: -

FINANCIAL BID

Purchase of “**UV Transilluminator**” for Anatomy Department at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	HSN Code	Make/ Brand/ Model	Unit Price in ₹.	GST Per Unit @	Total Unit Price with Tax in ₹.	Total Amount in ₹.
1.	<p>Ultra-violet (UV) transilluminator equipment to be used for visualization of DNAs and proteins agarose and polyacrylamide gels after electrophoresis.</p> <ul style="list-style-type: none">• The system should be extremely protected and covered from all sides.• Should have slot to place the gel for illumination and use hands from both the side to adjust the gel.• System should have provision to use the camera.• System should have metal body cover to limit or avoid direct exposure with the radiation.• The Transmission Size(W*L) should be not less than 200*150mm• The Transmission UV Wavelength of the system should be 302nm• Reflectance UV Wavelength should be 254, 365nm• The Transmission UV Lamp Power should be 302nm(8W)• The Reflectance UV Lamp Power should be 254nm(11W), 365nm(11W)• The system should have operating port for cutting the gel.• System should do viewing without requiring a dark room.• UV filter should be of Glass Quartz ultraviolet lamp.• The lamp should have a minimum self-life of 5 years and uniformity of the light should not vary with time.• The required Power Supply to run the system should be AC110/220V±10%, 50/60Hz• The dimension of the system not be more than (W*D*H) 435*295*490 mm	01 Nos.						

Sign of bidder :-
Date :-
Name of the bidder :-
Seal of the bidder :-