



No.: AIIMS/R/CS/Patho/20/26/LPC/A/

Date: 03.08.2020

QUOTATION NOTICE

Inviting quotations for purchase of “**Horizontal Gel Electrophoresis**” for Department of Pathology & Lab Medicine at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest/Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2nd Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **10.08.2020** before **3:00 pm**. The quotations will be opened on the same day at **3.30 pm**.

Terms & Condition:

1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed "**Technical Bid for Horizontal Gel Electrophoresis**" and "**Financial Bid for Horizontal Gel Electrophoresis**". Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as "**Quotation for supply of Horizontal Gel Electrophoresis**".
3. **All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.**
4. Rate should be mentioned in words & figure both.
5. Taxes/GST, if any (Kindly mention in above table)
6. Delivery schedule: within 15 days from the date of issue of P.O.
7. No additional documents related to this NIQ will be entertained after opening of NIQ.
8. Price should be FOR Destination basis (i.e. Department of Pathology, AIIMS Raipur).
9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
10. Quotation No/Name and due date of opening must be written on top of envelop.
11. **Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.**
12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned.**
13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
14. The GST registration details may be furnished.
15. RTGS details required for payment purpose.
16. 100% payment against receipt and acceptance of material.
17. No part supply or part payment will be entertained.
18. Validity of offer should not be less than 90 days.
19. Supply, installation and commissioning will be done by firm (if applicable).
20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.
22. Due of COVID-19 soft copy of quotation notice is acceptable on the company letterhead with seal and signature in storesofficer.cp@aiimsraipur.edu.in mail id up to 10.08.2020 before 3:00 pm.
23. The product should be of superior quality and highly standard.

Stores Officer,
AIIMS, Raipur (C.G.)



ANNEXURE-I

Ref. No. & Date: -
Quotation Notice No.: -
Brand/Make Name:-
Model No.:-
Warranty-

TECHNICAL BID

Purchase of “**Horizontal Gel Electrophoresis**” for Department of Pathology & Lab Medicine at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./Unit	Offered Specification	Specify Technically Complied/Not Complied / if any Deviations
1.	Wide horizontal electrophoresis system should include 15 & 20 – well combs, gel caster, 15x10cm (WxL) UV transparent tray, basic power supply unit.			
2.	The system should be made up of clear plastic construction for easy sample visualization.			
3.	The system should have UV – transparent gel trays with fluorescent ruler.			
4.	The system should have combs to fit every need (multichannel pipette compatible fixed height drop in combs, adjustable height combs & preparative combs).			
5.	The system should have easy – to – remove & replace electrode assemblies.			
6.	The system should have leak- proof design to prevent buffer leaks from the base.			
7.	The system should have arrow on the side of the base to indicate the direction of the run, ensuring proper gel orientation.			
8.	The system should have color-coded, labelled electrodes labelled base guarantee correct positioning of the lid on the base.			
9.	The system should have longer tabs on the base to prevent incorrect lid positioning & enable easy removal of the lid, reducing buffer spillage.			
10.	The system should have UV transparent gel trays, combs & other accessories.			
11.	Samples throughput of the system should be 1-120 approximately.			
12.	Base buffer volume capacity should be approx. 1.0 litre.			
13.	Cell size should be approx. 18x 40x 9.5 cm (WxLxH).			
14.	Distance to electrode should be around 30 cm			
15.	Bromophenol blue dye migration rate should be around 3.0 cm/hr (at 75V).			
16.	The system should be supplied with basic power supply (10-300V, 4-400mA, 75W, 4 Output jacks, Safety features- Overload/Short circuit detection, overvoltage protection, no load detection, sudden load change detection).			
17.	Power pac should have option for automatic recovery after power failure.			
18.	Power Pac should have option for constant current and constant voltage.			
19.	Power Pac Should meet EN-61010/ CE.			
20.	Support for induction and follow up training of technical staff, on –site standardization and troubleshooting of procedures/test to be provided by the company			
21.	After sale, service should be available promptly.			
22.	Should be CE or FDA or BIS approved product.			
23.	Warranty 12 Months for the electrophoresis unit and 24 Months for the Power Pac power supply			

Sign of bidder :-
Date :-
Name of the bidder :-
Seal of the bidder :-



ANNEXURE-II

Ref. No. & Date: -

Quotation Notice No.: -

FINANCIAL BID

Purchase of “**Horizontal Gel Electrophoresis**” for Department of Pathology & Lab Medicine at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./Unit	Offered Specification	Specify Technically Complied/Not Complied / if any Deviations
1.	Wide horizontal electrophoresis system should include 15 & 20 – well combs, gel caster, 15x10cm (WxL) UV transparent tray, basic power supply unit.			
2.	The system should be made up of clear plastic construction for easy sample visualization.			
3.	The system should have UV – transparent gel trays with fluorescent ruler.			
4.	The system should have combs to fit every need (multichannel pipette compatible fixed height drop in combs, adjustable height combs & preparative combs).			
5.	The system should have easy – to – remove & replace electrode assemblies.			
6.	The system should have leak- proof design to prevent buffer leaks from the base.			
7.	The system should have arrow on the side of the base to indicate the direction of the run, ensuring proper gel orientation.			
8.	The system should have color-coded, labelled electrodes labelled base guarantee correct positioning of the lid on the base.			
9.	The system should have longer tabs on the base to prevent incorrect lid positioning & enable easy removal of the lid, reducing buffer spillage.			
10.	The system should have UV transparent gel trays, combs & other accessories.			
11.	Samples throughput of the system should be 1-120 approximately.			
12.	Base buffer volume capacity should be approx. 1.0 litre.			
13.	Cell size should be approx. 18x 40x 9.5 cm (WxLxH).			
14.	Distance to electrode should be around 30 cm			
15.	Bromophenol blue dye migration rate should be around 3.0 cm/hr (at 75V).			
16.	The system should be supplied with basic power supply (10-300V, 4-400mA, 75W, 4 Output jacks, Safety features- Overload/Short circuit detection, overvoltage protection, no load detection, sudden load change detection).			
17.	Power pac should have option for automatic recovery after power failure.			
18.	Power Pac should have option for constant current and constant voltage.			
19.	Power Pac Should meet EN-61010/ CE.			
20.	Support for induction and follow up training of technical staff, on –site standardization and troubleshooting of procedures/test to be provided by the company			
21.	After sale, service should be available promptly.			
22.	Should be CE or FDA or BIS approved product.			
23.	Warranty 12 Months for the electrophoresis unit and 24 Months for the Power Pac power supply			

Sign of bidder :-
Date :-
Name of the bidder :-
Seal of the bidder :-