अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)



Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

No.: AIIMS/R/CS/Nephro/589/20/LPC

## **QUOTATION NOTICE**

Date: 22.12.2020

Inviting quotations for purchase of "DIALYZER WASHING SINK" for Nephrology Department at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, Ground Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **30.12.2020** before **3:00** pm.

### Terms & Condition:

- 1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
- 2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed "Technical Bid for DIALYZER WASHING SINK" and "Financial Bid for DIALYZER WASHING SINK". Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as "Quotation for supply of DIALYZER WASHING SINK" for Nephrology Department.
- **3.** All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.
- 4. Rate should be mentioned in words & figure both.
- 5. Taxes/GST, if any (Kindly mention in table)
- 6. Delivery schedule: within 15 days from the date of issue of P.O.
- 7. No additional documents related to this NIQ will be entertained after opening of NIQ.
- 8. Price should be FOR Destination basis (i.e. Neurosurgery Department, AIIMS Raipur).
- 9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
- 10. Quotation No/Name and due date of opening must be written on top of envelop.
- 11. Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.
- 12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned**.
- 13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following Anti Profiteering Clause of GST.

"Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices".

- 14. The GST registration details may be furnished.
- 15. RTGS details required for payment purpose.
- 16. 100% payment against receipt and acceptance of material.
- 17. No part supply or part payment will be entertained.
- 18. Validity of offer should not be less than 90 days.
- 19. Supply, installation and commissioning will be done by firm (if applicable).
- 20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.
- 22. Product should be of highly standard and superior quality.



# अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh) <sup>Tatibandh, GE Road, Raipur-492 099 (CG)</sup> <u>www.aiimsraipur.edu.in</u>

#### ANNEXURE-I

Ref. No. & Date: -Quotation Notice No.: -Brand/Make Name:-Model No.:-Warranty-

#### TECHNICAL BID

Purchase of "DIALYZER WASHING SINK" for Nephrology Department at AIIMS Raipur. Details of item are given as under:

S.	Description of item(s)	Qty/	Total	Offered	Specify Technically Complied /
<u>N.</u>		Set	Qty	Specification	Not Complied/ if any Deviations
1.	<b>Dialyzer washing sink</b> Fully stainless steel should made up of good grade of steel to prevent contamination dialyzer washing sink.	01	01		
	-needed to be filled on the wall				
	-sinks used for reprocessing should have a depth of at least 45cm with a drainage				
	-steel mesh at a depth of around 20cm to prevent the dialyzer and tubing resting in the effluent				
	-32 mm stainless steel waste coupling, ss mesh tray for keeping the dialyzer.				
	-(filter),pressure regulating valve with gauge,dimention 93''lx24.5''wx33.25''h inches-the hospital pattern (htm64)through has fully radius internal corners for ease of cleanning and hygiene.				
	-amt dt01 tap (6nos) with nozzles for various tube sizes, water gun with 8 different tips				
	-mannufacture should experience of supplying dialysis related equipments.				
	installation work will be done by the vendor				
	4 washing ports will be provided and installed by the vendor				

Sign of bidder	:-
Date	:-
Name of the bidder	:-
Seal of the bidder	:-



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh) <sub>Tatibandh, GE Road, Raipur-492 099 (CG)</sub>

www.aiimsraipur.edu.in

## ANNEXURE-II

Ref. No. & Date: -Quotation Notice No.: -

## FINANCIAL BID

Purchase of "DIALYZER WASHING SINK" for Nephrology Department at AIIMS Raipur. Details of item are given as under:

S. N.	<b>Description of item(s)</b>	Total Qty.	HSN Code	Make/ Brand/	Unit Price in	GST Per Unit @	Total Unit Price with	Total Amount in ₹.
		<b>C</b> -J-		Model	₹.	· · · · · · ·	Tax in ₹.	
1.	<b>Dialyzer washing sink</b> Fully stainless steel should made up of good grade of steel to prevent contamination dialyzer washing sink. -needed to be filled on the wall	01 Set						
	-sinks used for reprocessing should have a depth of at least 45cm with a drainage -steel mesh at a depth of around 20cm to							
	prevent the dialyzer and tubing resting in the effluent							
	-32 mm stainless steel waste coupling, ss mesh tray for keeping the dialyzer.							
	-(filter), pressure regulating valve with gauge, dimention 93''lx24.5''wx33.25''h inches-the hospital pattern (htm64) through has fully radius internal corners for ease of cleanning and hygiene.							
	-amt dt01 tap (6nos) with nozzles for various tube sizes, water gun with 8 different tips							
	-mannufacture should experience of supplying dialysis related equipments.							
	installation work will be done by the vendor							
	4 washing ports will be provided and installed by the vendor							

Sign of bidder	:-
Date	:-
Name of the bidder	:-
Seal of the bidder	:-