



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
**Tatibandh, GE Road,**  
**Raipur-492 099 (CG)**  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

AIIMS/R/HS/NEPH/676-678/155/2020

Date: 28/12/2020

**Inviting Quotations for Purchase of Medical Equipment & tools (Scrub Station and Hand wash Station) required for Nephrology Department at AIIMS Raipur.**

**NOTICE INVITING QUOTATION**

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Medical Equipment & tools (Scrub Station and Hand wash Station) required for Nephrology Department, AIIMS Raipur**. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer – Hospital, Room No. – 146, C C1 Block, Gate No. 1, AIIMS Raipur or through email at [quotations.hs@aiimsraipur.edu.in](mailto:quotations.hs@aiimsraipur.edu.in)** to 04/01/2021 before 03:00 pm. The online Bids can be submitted through CPP portal also. The quotations/Bids will be opened on 05/01/2021 at 03:30 pm. Details of item are given as under:-

Sn o.	Specification	Quantity Required	Unit	HSN Code	Brand	Unit Rate	GST			Amount
							SGST	IGST	CGST	
1.	Hand Wash Station	01	Nos							
2.	Scrub Station	01	Nos							
	Specification as per Annexure - I									
	Grand Total									

**Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.

4. Products are certified from ISI/ISO/CE/GMP/BIS as applicable, the Certificate to this effect should be attached.
5. Supply should be done within 15 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from concerned department.
- 8. Quotation Name/No. and due date of opening must be mentioned on top of envelopes.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms.  
The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
11. All other terms & condition as per GFR 2017.
12. Material to be delivered at **Nephrology Department, AIIMS Raipur.**
- 13. Validity of the quotation should be 90 days from the date of opening.**
- 14. Bidders should submit their email id, contact details with GST registration.**
- 15. The firm may participate through offline mode or online mode, the L1 will be selected among all the bidders irrespective of mode of participation.**

**Stores Officer (H)**  
**AIIMS, Raipur (C.G)**

## **Annexure-I**

### **Specification for Hand Wash Sink and related Accessories for Hemodialysis Unit:**

#### **1. Hand wash unit**

- The HWS unit should be made out of 18 /20 G SS 304 Q sheets duly matt finished.
- The wash station must be provided with auto taps, auto stainless steel liquid soap, IPA (sanitizer) dispenser & Hand dryer.
- A front door needs to be provided below the sink for maintenance.
- The unit should be either free standing or wall mounted (preferably).
- The water supply pipe will be of 0.5” and drainage pipe will be of 1.5” UPVC .
- Overall Dimension 43” x 18” x 34 + 24” H, Sink size – 530 x 325 x 200 mm D- 2, Auto taps – 2 nos. Soap dispenser – 1 no., IPA dispenser – 1 no., Hand dryer –1 no.

#### **2. SS Hand Dryer**

- Body Material: Stainless Steel (304G)
- Power: 1800 W
- Power Supply: AC 220V-50Hz
- Air Speed : 95 m/s
- RPM : 25000
- Drying Time : 7-10 seconds
- Noise Level: 75 db
- Dimension : 240\*174\*270 mm

#### **3. SS Soap & Ipa Dispenser**

- Body Material : Stainless Steel (304G)
- Capacity : 900ml
- Detection Zone : 10-15cms
- Drop : 1ml
- Dimension : 110\*275\*106 mm
- Battery : 6\*AA Duracell / Alkaline Battery
- Suitable For Alcohol Gel, Liquid Alcohol, Liquid Soap, Hand Sanitizers, Sterilised Iodine & Detergents

#### **4. Auto Sensor Taps**

- Power: DC 6V-220V/50Hz (AC)
- Battery : 4x AA Duracell / Alkaline Batteries
- Power Consumption : 0.5 mw (DC), <2 W (AC)
- Detection Zone : 12-18 cm
- Water Pressure : 0.05 – 0.6 Mpa
- Water Temperature : 0 – 60° C
- Degree of Protection : IP 56
- Available in AC & DC

#### **5. Installation and at least one year warranty is required to be provided by the vendor.**

### **Specification for Scrub Sinks for Haemodialysis Unit:**

- SS 304 surgical scrub sink (Floor Mounted ) Operation theatre fabricated from 18 Gauge (1.2 mm thick)
- SS 304 type seamless welded, polished to mat finish including connecting to drainage and water points and making good cuttings walls, Soap Dispenser, Instrument Tray.
- Size: L4' X W2' X H3'
- Spot 2 taps Connector; 4 No. ½'' male and Female Connecters and Water Joint Connector 7No. 12mm flexible PVC (pipe 08 mtr) anchor Fastener wall mounted 4 No. Drain Coupling flexible pipe 1 No. Foot paddle 2 No, and ½'' 24 Valve 2 No

**Note: This Sensor one year warranty.**

**Sensor – 2 Nos.; Soap Dispenser – 1 No.; Instrument tray – 1 No. (Including packing, transportation and installation).**