

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर(छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

Date: 28/05/2021

AIIMS/R/CS /Micro/20/200/B/LPC

विषय/Sub:Inviting Quotations for procurement of Table Top Micro Centrifuge with Rotor Head for Microbiology Department, at AIIMS Raipur.

## कोटेशन सूचना QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having GST Number and relevant documents for Table Top Micro Centrifuge with Rotor Head for Microbiology Department, at AIIMS Raipur. The quotation should be submitted to Medical College Building, Ground Floor, Besides SBI ATM Machine, Gate no.05 office of Store Officer up to 3:00pm on 01/06/2021. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

S.N	NAME	REQ	UNIT	Make &	UNIT	GST in	Unit Rate	TOTAL
0.		UIRE		HSN	RATE	%	with GST	PRICE
		MEN T		Code	IN Rs.			in RS.
1	Table Top Micro Centrifuge	03	Nos		NS.			
1	with Rotor Head	03	1103					
	Technical Specification:							
	General							
	i.The insturment should be	<u> </u>						
	ideal for routine application	<u> </u>						
	in clinical and research labs.							
	ii.The instrument should be	<u> </u>						
	supplied with 01 fixed angle	<u> </u>						
	rotor to support 1.5 ml micro-	<u> </u>						
	centrifuge tube.	<u> </u>						
	iii.The rotor should have 24 x	<u> </u>						
	1.5 /2 ml micro-centrifuge	<u> </u>						
	tube.	<u> </u>						
	iv.Max speed with fixed	<u> </u>						
	angle rotor: Maximum of							
	16000 rpm.	<u> </u>						
	v.Control System:	<u> </u>						
	Microprocessor based with	<u> </u>						
	LED display technology.	<u> </u>						
	vi.Instrument should have	<u> </u>						
	imbalance detector.							
	vii.The instrument should							
	have brushless induction	<u> </u>						
	motor.	<u> </u>						
	viii.The motor should be	<u> </u>						
	frequency drive.							
	ix.The instrument should							
	have dynamic brake.	 						
	x.Rotor should have safety							
	lid inter locking system to							
	prevent cover opening during							

centrifugation.				
xi.The instrument should				
have digital speed meter & 0-				
15 minutes' digital				
countdown timer.				
xii.Press switch for quick				
acceleration to full speed.				
Electrical :-				
i.Supply Voltage :230 ± 12V,				
AC, 50/60 Hz. Voltage				
ii.Plug to be adapted to meet				
the country requirements.				
iii.The line cord/power cord				
supplied with the equipment				
shall be of acceptable				
durability, length and current				
carrying complying with				
Indian Standards.				
Certification & Warranty:-				
i.The centrifuge should have				
a warranty for atleast 1 year.				
ii.The instrument should be				
IVD compliant, CE certified.				

## **Terms & Conditions.**

- 1. Rate should be mentioned in words & figures both.
- 2. **GST rates applicable** on your quoted item may please be confirmed.
- 3. Delivery Schedule- within 15 days from the date of issue of P.O.
- 4. No additional documents related to this NIQ will be entertained after NIQ opened.
- 5. Price should be F.O.R. for Destinatin basis (i.e.Concerned Department).
- 6. LD@0.5% of delayed supply per week or part week for delay of supply of material subject to maximum upto 10% of delayed supply should be deducted.
- 7. Quotation No/Name and Due date of opening must be written on top of the envelop.
- 8. The bidder participating in the LPC may have local regsitered office in chhattisgarh.
- 9. The GST registration details may please be furnished.
- 10. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profitering Clause of GST. "Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices"
- 11. Brand & Make should be clearly mentioned in offers
- 12. AIIMS Raipur reserved the rights to place order for full or part quantity to one or more firm.
- 13. Validity of offer should not be less than 90 days.
- 14. Supply, Installation and Commissioning will be done by firm (if applicable).
- 15. The quantity shown in above column are totally tentative, it may be increase and decrease at the time of placement of purchase order.
- 16. Due of COVID-19 soft copy of quotaiton notice is acceptable in storesofficer.cp@aiimsraipur.edu.in mail-id.

Stores Officer AIIMS Raipur (C.G.)

## **Vendor Details**

Name	
Aadhaar No. (if any)	
PAN	
GST.NO.	
Address	
City	
State	
Pincode	
Mobile No.	
Phone No.	
E-mail	
Bank Name	
Bank A/c No.	