



AIIMS/R/CS /Micro/22/35/LPC

Date: 28/02/2022

विषय/Sub:Inviting Quotations for procurement of Micropipette Single Channel Variable Volume (100-1000UL) for Microbiology Department, at AIIMS Raipur.

कोटेशन सूचना QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having GST Number and relevant documents for Micropipette Single Channel Variable Volume (100-1000UL) for Microbiology Department, at AIIMS Raipur. The quotation should be submitted to Medical College Building, Ground Floor, Besides SBI ATM Machine, Gate no.05 office of Store Officer up to 3:00pm on 07/03/2022. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

S.N o.	NAME	REQUIREMENT	UNIT	Make & HSN Code	UNIT RATE IN Rs.	GST in %	Unit Rate with GST	TOTAL PRICE in RS.
1	Micropipette Single Channel Volume: 100 - 1000 ul	02	Nos					

Specifications for Adjustable Single Channel Micropipette.

I.General specification for adjustable micropipette

- 1.Micropipette should be adjustable with ultra- light weight.
- 2.Micropipette should be fully autoclavable.
- 3.Adjustment opening for adjusting pipettes to a specific liquid and volume.
- 4.Piston System: Ultra-light system and should be made of fortom.
- 5.It should be highly resistant to heat, acids and alkali. mildew, bleaches, aging, sunlight and abrasion.
- 6.No discoloration upon UV irradiation.
- 7.It should have button to control very low operating force, Color indication for pipette volume, Positioned for perfect Ergonomics.
- 8.Volume Display: Four digit display with magnifier and 2 Button operations.
- 9.Improved ergonomics to reduce stress without sacrificing tightness.
- 10.Ejector: Very low ejection force and positioned for perfect ergonomics.
- 11.Quick connection clip: Remove lower part easily.
- 12.Viable calibration seal to indicate factory calibration not changed.
- 13.Vendor should have NABL calibration facility, The proof for the same has to be enclosed along with the Tender.
- 14.Each micropipette should be supplied with factory calibration certificate.
- 15.Each Micropipette upto 1 ml volume should have autoclavable tip box with 96 tips.
- 16.Micropipette should be compatible with tips of any other manufacturer.
- 17.The item should be ISO 8655 and IVD compliant.

Technical Data:

a. 100-1000 μ L adjustable micropipette.

Volume range – 100 – 1000 μ L		
Volume	Systematic error (approx.)	Random error (apprx.)
100 μ L	$\pm 3.0 \mu$ L	$\pm 0.6 \mu$ L
500 μ L	$\pm 5.0 \mu$ L	$\pm 1 \mu$ L
1000 μ L	$\pm 6.0 \mu$ L	$\pm 2 \mu$ L

Note:1.Quotation without technical catalogue will not be considered.

2.Items should be available for demonstration.

3.Should come with three years warranty from the date of supply.

4.Supplier should ensure the service and supply of spares beyond the warranty period.

Terms & Conditions.

1. Rate should be mentioned in words & figures both.
2. **GST rates applicable** on your quoted item may please be confirmed.
3. Delivery Schedule- within 15 days from the date of issue of P.O.
4. No additional documents related to this NIQ will be entertained after NIQ opened.
5. Price should be F.O.R. for Destination basis (i.e.Concerned Department).
6. LD@0.5% of delayed supply per week or part week for delay of supply of material subject to maximum upto 10% of delayed supply should be deducted.
7. Quotation No/Name and Due date of opening must be written on top of the envelope.
8. The bidder participating in the LPC may have local registered office in Chhattisgarh.
9. **The GST registration details may please be furnished.**
10. Please confirm if there any change (Upward/Reduction) in your **Basic Price** structure. And you are also requested to pass the Input Credit as per the following **Anti Profiteering Clause** of GST.**“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”**
11. Brand & Make should be clearly mentioned in offers
12. AIIMS Raipur reserved the rights to place order for full or part quantity to one or more firm.
13. Validity of offer should not be less than 90 days.
14. Supply,Installation and Commissioning will be done by firm (if applicable).
15. The quantity shown in above column are totally tentative, it may be increase and decrease at the time of placement of purchase order.

Sr.Procu.Cum.Stores Officer
AIIMS Raipur (C.G.)

Vendor Details

Name	
Aadhaar No. (if any)	
PAN	
GST.NO.	
Address	
City	
State	
Pincode	
Mobile No.	
Phone No.	
E-mail	
Bank Name	
Bank A/c No.	