



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/R/HS/NEPH/993/155/2022

Date: 01/04/2022

NOTICE INVITING QUOTATION

Inviting Quotations for Purchase of Paediatric Low flux Dialyzer 0.7 - 0.9m² required for Haemodialysis unit at Nephrology Department at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors/traders having GST and relevant documents for Purchase of **Paediatric Low flux Dialyzer 0.7 – 0.9 m² required for Haemodialysis unit at Nephrology Department, AIIMS Raipur.** The quotation with copy of certificate of GST & other documents should be submitted to office of **Stores Office – Hospital, Room No. –329, C C1 Block, Gate No. 1, AIIMS Raipur up to 05/04/2022 before 03:00 pm.** The quotations/Bids will be opened on the same day i.e. on **05/04/2022 at 03:30pm.** Details of item are given as under:-

Sno	Specification	Quantity Required	Unit	HSN Code	Brand	Unit Rate	GST			Amount
							SGST	IGST	CGST	
1.	Paediatric Low flux Dialyzer 0.7 – 0.9 m ² Specification as per Annexure - I	140	Nos.							
Grand Total										

Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Supply should be done within 15 days after Placement of PO.
5. Price should be FOR Destination basis (i.e. concerned department).
6. 100% Payment will be released after certification from concerned department.
7. **Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
8. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.

9. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms.
The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
10. All other terms & condition as per GFR 2017.
11. Material to be delivered at **Nephrology Department, AIIMS Raipur.**
12. **Validity of the quotation should be 90 days from the date of opening.**
13. **Bidders should submit their email id, contact details with GST registration.**
14. **AIIMS Raipur reserves the right to ask the tenderers for arranging demonstration of their samples for feel & amp; finish for which rates have been quoted, to the concerned committee, if required.**

Stores Officer (H)
AIIMS, Raipur (C.G)

Annexure – I

Pediatric Low flux (8-10) dialyser of surface area 0.7 to 0.9 m²

1. Membrane material should be made up of Polysulfone/ Polynephron/ Polyflux (a blend of Polyarylethersulfone, Polyvinylpyrrolidone, and Polyamide) with a thickness of wall (membrane wall) 32 to 40 μm . Inner diameter of lumen to be limited to 190 to 220 μm range.
2. Sterilization method of dialyzer: inline steam/ beta or gamma ray (including dry gamma) with potting compound to be made up of only polyurethane and Housing material to be made up of only polypropylene/ polycarbonate.
3. Dialyzer has to be Biocompatible with high endotoxin retention capabilities with No chemical residuals and Low rinsing volumes.
4. Dialyzer should allow TMP 450 to 500 mm of hg for low flux, 450 to 550 mm of hg for middle flux and 550 to 650 mm of hg for high flux.
5. KUF (ml/h/mm of hg) low flux 8-10.
6. Dialyzers are to be produced to meet international safety regulations.