



NO. AIIMS/R/HS/2022/155/182

Date 25.11.2022

Sub: Inviting quotations for procurement of Laparoscopic Gastrointestinal Universal Stapler for department of Surgical Gastroenterology at AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/Distributors/Traders having GST No./relevant documents for procurement of Laparoscopic Gastrointestinal Universal Stapler for department of Surgical Gastroenterology at AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to **office of Stores Officer – Hospital, Room No. – 146, C-C1 Block, Gate No. 1, AIIMS Raipur up to 02/12/2022 before 03:00 pm**. The quotations will be opened on the same day at 03:30 pm. Details of items are as under :-

S.n क्र.स.	Name of the Item आइटम कानाम	HSN code एच.एस. एन.	Make/ Model	Maximum Order Qty. मात्रा	UNIT RATE IN Rs. इकाई दररुमें	GST जीएसटी	UNIT RATE with GST इकाई दर करसहित	Total amount in Rs. कुल मूल्य
1.	Laparoscopic Gastrointestinal Universal Stapler short – 6 cm shaft			03 unit				
2.	Laparoscopic Gastrointestinal Universal Stapler Long – 26 cm shaft			03 unit				

Specification of the products – attached at Annexure-A

Terms & Condition:

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Products are certified from ISI/ISO/CE/BIS as applicable, the Certificate to this effect should be attached.
5. Supply should be done within 15 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from concerned department.
8. **Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.

10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
11. All other terms & condition as per GFR 2017.
12. Material to be delivered at Surgical Gastroenterology department at AIIMS Raipur.
13. Validity of the quotation should be 90 days from the date of opening.
14. Firm to submit documentary evidence in support of their claim for GST at the time of submission of bills.

Stores Officer (H)
AIIMS, Raipur (C.G)

Annexure-A

Specification for:

A) Laparoscopic Gastrointestinal Universal Stapler Short - 6 cm shaft

Quantity: 3 units

1. Laparoscopic Gastrointestinal Stapler should have facility for interchanging cartridge lengths of 30mm, 45mm and 60mm in the same stapler with Tissue Gap Control Mechanism
2. there should be 9 Degree Articulation with 5 point Articulation on Each Side;
3. The handle when squeezed should move only 15mm at a time, ensuring precise staple formation even at the distal end
4. The firing sequence of the stapler to be made possible only by activating a push button
5. Having for at least 25 firings with a facility for 360° rotation
6. Should also facilitate a 22 degree and 45 degree articulation on either side of the stapler with Roticulating cartridges.
7. Shaft length of 6 cm - suitable for open/laparoscopic surgeries
8. CE, USFDA and ISO certified

B) Laparoscopic Gastrointestinal Universal Stapler Long - 26 cm shaft

Quantity: 3 units

1. Laparoscopic Gastrointestinal Stapler should have facility for interchanging cartridge lengths of 30mm, 45mm and 60mm in the same stapler with Tissue Gap Control Mechanism
2. there should be 9 Degree Articulation with 5 point Articulation on Each Side;
3. The handle when squeezed should move only 15mm at a time, ensuring precise staple formation even at the distal end
4. The firing sequence of the stapler to be made possible only by activating a push button
5. Having for at least 25 firings with a facility for 360° rotation
6. Should also facilitate a 22 degree and 45 degree articulation on either side of the stapler with Roticulating cartridges.
7. Shaft length of 26 cm - suitable for laparoscopic surgeries
8. CE, USFDA and ISO certified