



NO. AIIMS/R/HS/2022/155/183

Date 25.11.2022

Sub: Inviting quotations for procurement of Surgical Instrument for department of Surgical Gastroenterology at AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/Distributors/Traders having GST No./relevant documents for procurement of Surgical Instrument for department of Surgical Gastroenterology at AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to **office of Stores Officer – Hospital, Room No. – 146, C-C1 Block, Gate No. 1, AIIMS Raipur up to 02/12/2022 before 03:00 pm**. The quotations will be opened on the same day at 03:30pm. Details of items are as under :-

S.n क्र.स.	Name of the Item आइटम कानाम	HSN code एच.एस. एन.	Make/ Model	Maximum Order Qty. मात्रा	UNIT RATE IN Rs. इकाई दररुमें	GST जीएसटी	UNIT RATE with GST इकाई दर करसहित	Total amount in Rs. कुल मूल्य
1.	Trocar with sleeve (Optiport)-5 mm			06 pcs.				
2.	Trocar with sleeve (Optiport)-12 mm			09 pcs.				
3.	Trocar with sleeve (Optiport)-15 mm			03 pcs.				
4.	Contour curved cutter-40 mm			01 pcs.				
Specification of the products – attached at Annexure-A								

Terms & Condition:

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Products are certified from ISI/ISO/CE/BIS as applicable, the Certificate to this effect should be attached.
5. Supply should be done within 15 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from concerned department.

- 8. Quotation Name/No. and due date of opening must be mentioned on top of envelopes.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
11. All other terms & condition as per GFR 2017.
12. Material to be delivered at Surgical Gastroenterology department at AIIMS Raipur.
13. Validity of the quotation should be 90 days from the date of opening.
14. Firm to submit documentary evidence in support of their claim for GST at the time of submission of bills.

**Stores Officer (H)
AIIMS, Raipur (C.G)**

Annexure-A

Specification for:

A) Trocar with Sleeve (optiport) - 15 mm:

Quantity: 3 units

1. Optically guided bladeless trocar 15 mm diameter with bilateral tissue separators
2. length 100mm
3. optical tip to eliminate blind entry
4. clear ribbed cannula
5. recessed stopcock valve
6. funnel shaped housing
7. duckbill secondary seal and integrated universal seal that eliminates the use of reducer.
8. USFDA certified

B) Trocar with Sleeve (optiport) - 12 mm:

Quantity: 9 units

1. Optically guided bladeless trocar 12 mm diameter with bilateral tissue separators
2. length 100mm
3. optical tip to eliminate blind entry
4. clear ribbed cannula
5. recessed stopcock valve
6. funnel shaped housing
7. duckbill secondary seal and integrated universal seal that eliminates the use of reducer.
8. USFDA certified

C) Trocar with Sleeve (optiport) - 5 mm:

Quantity: 6 units

1. Optically guided bladeless trocar 5 mm diameter with bilateral tissue separators
2. length 100mm
3. optical tip to eliminate blind entry
4. clear ribbed cannula
5. recessed stopcock valve
6. funnel shaped housing
7. duckbill secondary seal and integrated universal seal that eliminates the use of reducer.
8. USFDA certified

D) Contour Curved cutter 40 mm with Blue reload:

Quantity: 1 unit

1. 40 mm linear cutter with simultaneous cutting and stapling
2. unique curved head design for lower pelvic access
3. Max firings: 6
4. Should have manual pin placement option
5. reloadable cartridges facility should be available
6. USFDA approved