



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

AIIMS/R/HS/SUR-GAS/119/2023/PAC

Date: 06/04/2023

**Inviting Quotations for Purchase of Laparoscopic powered endocutter stapler and cartridge for Surgical Gastroenterology Department at AIIMS Raipur.**

**NOTICE INVITING QUOTATION**

Sealed quotations are invited from intending registered Stockist / Distributors/traders having GST and relevant documents for Purchase of **Laparoscopic powered endocutter stapler and cartridge for Surgical Gastroenterology Department** at AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted **through CPPP only upto 12/04/2023 before 03:00 pm**. The quotations/Bids will be opened on the next day i.e. on **13/04/2023 at 03:30pm**. Details of item are given as under:-

Sno.	Specification	Quantity Required	Unit	HSN Code	Brand	Unit Rate	GST			Amount
	Laparoscopic powered Endocutter stapler and cartridge						SGST	IGST	CGST	
1.	Battery Operated 60mm Articulating Endo Cutter	01	Nos.							
2.	Battery Operated 45mm Articulating Endo Cutter	01	Nos.							
3.	Endoscopic linear cutter cartridge for regular thick tissues - 60 mm (gold)	05	Nos.							
4.	Endoscopic linear cutter cartridge for thick tissues - 60 mm (green)	05	Nos.							
	Specification as per Annexure - A									
	The Bidders must submit the Purchase order copies of the same item supplied to other Central Government institutes or AIIMS.									
	Grand Total									

## **Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Supply should be done within 15 days after Placement of PO.
5. Price should be FOR Destination basis (i.e. concerned department).
6. 100% Payment will be released after certification from concerned department.
7. **Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
8. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
9. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
10. All other terms & condition as per GFR 2017.
11. Material to be delivered at **Surgical Gastroentrology Department, AIIMS Raipur.**
12. **Validity of the quotation should be 90 days from the date of opening.**
13. **Bidders should submit their email id, contact details with GST registration.**
14. **The Bidders must submit the Purchase order copies of the same item supplied to other central government institutes or AIIMS.**

**Stores Officer (H)**

**AIIMS, Raipur (C.G)**

## **Annexure -A**

### **Technical Specification:**

#### **Laparoscopic powered Endocutter stapler and cartridge - 1 set**

serial no	item	description	Quantity needed
1	Battery Operated 60mm Articulating Endo Cutter	It should be Battery Operated 60mm Articulating Endo Cutter with a disposable battery pack; should have enhanced distal tip stability while firing; should have closed channel in the cartridge jaw for better stability during firing; 360 degree rotation shaft and one handed natural articulation up to 45 degrees; should have Precision Machined Anvil to deliver initial, system wide compression; should have Wide Proximal to Distal Jaw Aperture (proximal 8mm , Distal 22mm); should have 3 Point Gap Control for alignment and calibration throughout the 60 mm staple line; should have Knife direction/reverse control to discontinue the firing and return the knife; should allow interchangeable 6 row cartridge options of closed staple height of 1.0, 1.5, 1.8, 2.0 and 2.3 mm and all fits down to 12mm trocar sleeve; 340 mm shaft length. (BIS/USFDA/CE certified as applicable)	1
2	Battery Operated 45mm Articulating Endo Cutter	It should be Battery Operated 45mm Articulating Endo Cutter with a disposable battery pack; should have enhanced distal tip stability while firing; should have closed channel in the cartridge jaw for better stability during firing; 360 degree rotation shaft and one handed natural articulation up to 45 degrees; should have Precision Machined Anvil to deliver initial, system wide compression; should have Wide Proximal to Distal Jaw Aperture (proximal 8mm , Distal 22mm); should have 3 Point Gap Control for alignment and calibration throughout the 60 mm staple line; should have Knife direction/reverse control to discontinue the firing and return the knife; should allow interchangeable 6 row cartridge options of closed staple height of 1.0, 1.5, 1.8, 2.0 and 2.3 mm and all fits down to 12mm trocar sleeve; 340 mm shaft length. (BIS/USFDA/CE certified as applicable)	1
3	Endoscopic linear cutter cartridge for regular thick tissues - 60 mm (gold)	Endoscopic linear cutter cartridge with pocket extension having 60mm staple line with 6 rows of staples, for regular thick tissues, cartridge having closed staple height of 1.8 mm compatible with the above mentioned Battery operated 60 mm articulating endo cutter. (BIS/USFDA/CE certified as applicable)	5
4	Endoscopic linear cutter cartridge for thick tissues - 60 mm (green)	Endoscopic linear cutter cartridge with pocket extension having 60mm staple line with 6 rows of staples, for thick tissues, cartridge having closed staple height of 2.0 mm compatible with the above mentioned Battery operated 60 mm articulating endo cutter. (BIS/USFDA/CE certified as applicable)	5

1. All the items should be from the same manufacturing company
2. Item wise selection as L-1 shall not be done

3. The bidder who has not quoted for all the instruments i.e. the entire set shall not be entertained and bid shall be rejected
4. Items must be USFDA, CE, ISO certified.

#Note 1: final quantity of cartridges may vary depending on the final rate by the bidding firm.