



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर, (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/R/HS/2026-27/Bio/Anti/155/ 837

Date: 07/05/2026

Sub: Various Consumables for Anti Phospholipids of Biochemistry Department at AIIMS Raipur.

NOTICE INVITING QUOTATION

Sealed quotations are invited from intending registered Stockist / Distributors/traders having GST and relevant documents for Purchase of: **Various Consumables for Anti Phospholipids of Biochemistry Department at AIIMS Raipur.** The quotation with copy of certificate of GST & other documents should be submitted to office of **Stores Office – Hospital, Hospital Administration Block 1st floor, Gate No. 1, AIIMS Raipur up to 14/05/2026 before 03:00 pm.** The quotations/Bids will be opened on the same day i.e. on **14/05/2026 at 03:30pm.** Details of item are given as under:-

Sno	Specification	Quantity Required	Unit	HSN Code	Brand	Unit Rate	GST			Amount
							SGST	IGST	CGST	
1.	Anti Cardiolipon ELISA IgM (Pack Size-96x01) 1.The test kit should be exclusively crafted for the quantitative (1 determination of human autoantibodies of the IgM class against cardiolipin with B2 GP1 as cofactor in serum or plasma for the diagnosis of Anti-phospholipid syndrome. 2. The test kit should utilize microtiter strips each with 8 break- off reagent wells coated with cardiolipin with 2 GPI as cofactor. 3. The test kit should contain 3 calibrators, control (positive and negative), enzyme conjugate, Sample buffer, Wash buffer, substrate solution and Stop solution, Test Instruction, and QC sheet. 4.The test kit lower detection limit should be 1.6 PL-IgM-U/ml 5.The test kit should have a linearity of 6 - 103 PL-IgM-U/ml 6.The test kit should be ready to use color coded reagents except for the enzyme conjugate and wash buffer solution. 7.The test kit should contain break-apart wells 8.The test kit should be automatable with minimum incubation time at RT	07								

	9.The test kit should be both CE & IVD certified.									
2.	<p>Anti Cardiolipon ELISA IgG (Pack Size-96x01)</p> <p>1.The test kit should be exclusively crafted for the quantitative... determination of human autoantibodies of the IgG class against cardiolipin with B2 GP1 as cofactor in serum or plasma for the diagnosis of Anti-phospholipid syndrome.</p> <p>2. The test kit should utilize microtiter strips each with 8 break- off reagent wells coated with Human B2-glycoprotein 1</p> <p>3. The test kit should contain 3 calibrators, control (positive and negative), enzyme conjugate, Sample buffer, Wash buffer, substrate solution and Stop solution, Test Instruction, and QC sheet.</p> <p>4.The test kit lower detection limit should be 0.8 PL-IgG-U/ml.</p> <p>5.The test kit should have a linearity of 4 PL-IgG-U/ml to 98 PL-IgG-U/ml.</p> <p>6. The test kit should have ready to use color coded reagents except for the enzyme conjugate and wash buffer solution.</p> <p>7. The test kit should contain break-apart wells</p> <p>8.The test kit should be automatable with minimum incubation time at RT</p> <p>9.The test kit should be both CE & IVD certified</p>	07								
3.	<p>Anti-Beta2 Glycoprotein 1 ELISA (IgM) (Pack Size-96x01)</p> <p>1. The test kit should be exclusively crafted for the quantitative determination of human autoantibodies of the IgM class against β2-glycoprotein 1 in serum or plasma for the diagnosis of Anti-phospholipid syndrome.</p> <p>2. The test kit should utilize microtiter strips each with 8 break-off reagent wells coated with Human B2-glycoprotein I</p> <p>3. The test kit should contain 3 calibrators, control (positive and negative), enzyme conjugate, Sample buffer, Wash buffer, substrate solution and Stop solution, Test Instruction, and QC sheet.</p> <p>4. The test kit lower detection limit should be 0.8 RU/ml.</p> <p>5. The test kit should have a linearity of 11-182 RU/ml</p> <p>6. The test kit should have ready to use color coded reagents except for the wash buffer solution.</p> <p>7. The test kit should contain break-</p>	07								

	<p>apart wells</p> <p>8. The test kit should be automatable with minimum incubation time at RT</p> <p>9. The test kit should be both CE & IVD certified</p>									
4.	<p>Anti-Beta2 Glycoprotein 1 ELISA (IgG) (Pack Size-96x01)</p> <p>1.The test kit should be exclusively crafted for the quantitative determination of human autoantibodies of the IgG class against β2-glycoprotein 1 in serum or plasma for the diagnosis of Anti-phospholipid syndrome.</p> <p>2. The test kit should utilize microtiter strips each with 8 break-off reagent wells coated with Human β2-glycoprotein 1</p> <p>3. The test kit should contain 3 calibrators, control (positive and negative), enzyme conjugate, Sample buffer, Wash buffer, substrate solution and Stop solution, Test Instruction, and QC sheet.</p> <p>4.The test kit lower detection limit should be 1.8 RU/ml.</p> <p>5. The test kit should have a linearity of 09-181 RU/ml</p> <p>6.The test kit should have ready to use color coded reagents except for the wash buffer solution.</p> <p>7. The test kit should contain break-apart wells</p> <p>8 The test kit should be automatable with minimum incubation time at RT</p> <p>9.The test kit should be both CE & IVD certified</p>	07								
Grand Total										

Note:

1. *Bidders are required to quote for all items. The L1 bidder will be determined based on the aggregate cost of all items quoted. The successful bidder shall carry out standardization of the parameters on the ELISA machine in the Department.*
2. *Upon successful standardization of parameters, the required kits shall be supplied, and payments will be processed upon successful standardization of the parameters.*

Terms & Condition:

1. **The interested agencies are required to submit the sealed technical bid and sealed financial bid separately. The bids in sealed Cover -I must comprise of “Technical Bid” and sealed Cover-II must comprise of “Financial Bid” Both these two cover must be placed in third sealed cover super-scribed with all the details mentioned.**
2. **Technical Bid must comprise of complete specification of item and other related document as asked in the NIQ.**
3. **Financial bids will be opened of only those bidders whose samples are technically qualified.**
4. **Rate should be mentioned in words & figures both.**
5. **GST rates applicable** on your quoted item may please be confirmed.

6. Delivery Schedule-

Item Details	To be delivered with In 45 days of PO released	To be delivered with In 90 days of PO released	To be delivered with In 180 days of PO released
Anti Cardiolipon ELISA IgM Pack Size-96x01	2 Kits	2 Kits	3 Kits
Anti Cardiolipon ELISA IgG Pack Size-96x01	2 Kits	2 Kits	3 Kits
Anti-Beta2 Glycoprotein 1 ELISA (IgM) (Pack Size-96x01)	2 Kits	2 Kits	3 Kits
Anti-Beta2 Glycoprotein 1 ELISA (IgG) (Pack Size-96x01)	2 Kits	2 Kits	3 Kits

7. No additional documents related to this NIQ will be entertained after NIQ opened.
8. Price should be F.O.R. for Destination basis (i.e. Concerned Department).
9. LD@0.5% of delayed supply per week or part week for delay of supply of material subject to maximum upto 10% of delayed supply should be deducted.
10. Quotation No/Name and Due date of opening must be written on top of the envelop.
11. **The GST registration details may please be furnished.**
12. Please confirm if there any change (Upward/Reduction) in your **Basic Price** structure. And you are also requested to pass the Input Credit as per the following **Anti Profiteering Clause** of GST. **“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”**
13. **Brand & Make should be clearly mentioned in offers.**
14. Technical Specification should be attached with rate Quotation.
15. AIIMS Raipur reserved the rights to place order for full or part quantity to one or more firm.
16. Material to be delivered at **Biochemistry Department, AIIMS Raipur.**
17. **Validity of the quotation should be 90 days from the date of opening.**
18. **Bidders should submit their email id, contact details.**
19. Supply, Installation and Commissioning will be done by firm (if applicable).
20. The quantity shown in above column are totally tentative, it may be increase and decrease at the time of placement of purchase order.
21. Specification should be enclosed with rate quotation.
22. **L1 will be decided Rate wise on FOR AIIMS basis.**

S/d

**Assistant Stores Officer
Procurement Lab
AIIMS, Raipur (C.G)**