



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)
storesofficer.hp@aiimsraipur.edu.in

NO. AIIMS/RPR/T&E/24200/2026/155/ 861

Date 07/07/2026

Sub: Inviting quotations for procurement of Laryngoscope Blade MCCOY type No. 3 & No. 4 for Trauma & Emergency department, AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/Distributors/Traders having GST No./relevant documents for supply of **Laryngoscope Blade MCCOY type No. 3 & No. 4 for Trauma & Emergency department, AIIMS Raipur**. The quotation with copy of certificate of GST & other documents should be submitted to **Gate No. 5, New Administrative Building A-Block 1st Floor, Hospital Store AIIMS Raipur up to 13/07/2026 before 12:00 pm**. The quotations will be opened on the same day at 12:30pm. Details of items are as under

S.n क्र.सं.	Name of the Item आइटम कानाम	Make/ Brand	HSN code एच.एस. एन.	Qty. मात्रा	UNIT RATE IN Rs. इकाई दर रु में	GST जीएसटी	UNIT RATE with GST इकाई दर कर सहित	Total amount in Rs. कुल मूल्य
1.	Laryngoscope Blade MCCOY type No. 3 & No. 4 No. 3 - 2Nos. No. 4 - 2Nos. Specification as per annexure - A		4 Nos.					

Technical Specification McCoy Type Laryngoscope Blade No. 3 & No. 4
Item: McCoy Type Laryngoscope Blade (Adult)
Sizes Required: No. 3 and No. 4

Technical Specifications

1. The blade shall be of McCoy (levering tip) type, designed to facilitate difficult airway management by providing elevation of the epiglottis through a hinged distal tip.
2. The blade shall be manufactured from high-quality medical-grade stainless steel, corrosion-resistant and suitable for repeated sterilization.
3. The blade shall have a smooth satin/matte finish to minimize glare during laryngoscopy.
4. The blade shall be available in Size No. 3 and Size No. 4 (adult).
5. The blade shall be compatible with standard ISO 7376 laryngoscope handles available in the institute.

6. The levering tip mechanism shall be smooth, robust, and ergonomically operated, providing controlled elevation of the distal tip without excessive force.
7. The blade shall incorporate a fiber-optic light transmission system with no exposed electrical contacts on the blade, providing bright and uniform illumination of the airway.
8. The fiber-optic bundle shall be removable or designed to facilitate effective cleaning and sterilization, where applicable.
9. The blade shall be fully autoclavable and capable of withstanding repeated steam sterilization cycles at temperatures up to 134°C without deterioration in performance.
10. The blade shall be free from sharp edges, burrs, or manufacturing defects and shall provide an atraumatic design suitable for routine clinical use.
11. The blade shall be latex-free and manufactured using biocompatible materials.
12. The device shall comply with applicable international quality and safety standards such as ISO 7376, ISO 13485, and shall be CE/BIS Certified/ US FDA approved/cleared, wherever applicable.
13. The product shall be supplied with the manufacturer's catalogue, user instructions, and cleaning/sterilization guidelines.
14. The manufacturer should have an established service and supply network in India for replacement parts and technical support.
15. The product shall carry a minimum warranty of one year against manufacturing defects.

The bidder shall submit:


1. Technical literature/brochure of the quoted model.
2. Compliance statement against each specification.
3. Valid CE Certificate/US FDA/BIS approval
4. ISO 13485 certification of the manufacturer.
5. Declaration of compatibility with standard ISO 7376 laryngoscope handles.
6. If found technically suitable bidder may be asked to
7. The purchaser reserves the right to ask the bidder to provide a live product demonstration of the quoted product during the technical evaluation process.
 - o The demonstration, if requested, shall be arranged by the bidder at no additional cost to the purchaser within the stipulated time.
 - o Failure to provide the demonstration or inability of the quoted product to meet the technical specifications during the demonstration may lead to rejection of the bid.


Store Officer (H)
AIIMS, Raipur (C.G)

पवन कुमार
Pawan Kumar
संस्कार अधिकारी (अस्प.)/Stores Officer (Hosp.)
अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.प्र.)
All India Institute of Medical Sciences, Raipur (C.G.)

Terms & Condition:

1. The sealed envelope should bear the quotation number and the subject of quotation.
2. Supply should be made within 15 days from the date of purchase order.
3. Price should be inclusive of freight on road charges.
4. Item was selection as L-1 shall not be done, the selection of L-1 shall be made on the basis of aggregate total value.
5. Validity of the quotation should be 60 days from the date of opening and validity period must be clearly mentioned in the quotation.
6. Sample to be submitted as and when asked for.
7. Selected Supplier must have a Valid License for supply of the demanded items.
8. Firm to submit documentary evidence in support of GST claim at the time of submission of bills.
9. Short-life items (shelf-life \leq 18 months) shall not have exceeded 5/6th of their total shelf life at the time of supply. For imported drugs, shelf life at the time of supply shall not be less than 50%.
10. For items not covered under clause (12), the supplied stock shall not be older than one year from the date of manufacture at the time of supply.
11. If supplied items are not utilized before expiry, the supplier shall undertake to replace them with fresh stock as and when required.
12. Supplier shall arrange for free replacement of any quantity that deteriorates in potency, strength, or approaches expiry (or has expired) before the expiry date marked on the labels.
13. Quotation should be submitted only during Working Days & Office hours:
14. Monday to Friday: 9 AM – 5 PM
15. Saturday: 9 AM – 1 PM
16. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
17. 100% Payment will be released after certification from concerned department.
18. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
19. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantities
20. All other terms & condition as per GFR 2017.
21. Material to be delivered at Trauma & Emergency, AIIMS Raipur.


Store Officer (H)
AIIMS, Raipur (C.G.)
Pawan Kumar
भंडार अधिकारी (अस्प.)/Stores Officer (Hosp.)
राष्ट्रीय भारतीय आयुर्विज्ञान संस्थान रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (C.G.)