



AIIMS/R/CS/Neuro/ Cautery/22/17/OT /corrigendum/

दिनांक:- 01.06.2018

Corrigendum-3

Tender ID No.: 2018_IMSRP_325057_1

With Reference to above tender ID No., the following amendment is being issued:-

Page No/ Clause no/ point no.	Existing sentence	To be read as/Addendum
Page No. 12, point f of Technical bid of tender document	Should be submit Indian Standard Certificate (BIS)/European CE certificate and USFDA certificate as mentioned in specification at Annexure-I & II.	Should submit BIS/ISO/European CE/USFDA certificate as mentioned in specification at Annexure-I & II
Page no. 12, 17, 18, 20, 21 & 22	All Units/Items and accessories should be Indian Standard (BIS)/European CE and USFDA approved	All Units/Items and accessories should be BIS/ISO/European CE/USFDA approved and should be from same manufacturer.
Page no. 17 & 21 point D	If any manufacturer has only bipolar unit with integrated irrigation then a separate Monopolar unit with all accessories and customized trolley, all of which is Indian Standard Certified (BIS)/USFDA & ECE certified should be supplied along with.	If any manufacturer has only bipolar unit with integrated irrigation then a separate Monopolar unit with all accessories and customized trolley, all of which is BIS/ISO/European CE/USFDA certified should be supplied along with.
Page no. 20 point No. 10 g	The unit should be Indian standard(BIS)/US FDA approved and should compile all IEC norms	The unit should be BIS/ISO/European CE/USFDA approved and should compile all BIS/ISO/European CE/USFDA IEC norms
Page No. 17 & 21	Price of essential consumables should be fixed for next 5 years.	No consumables required
Page No. 17 & 21	User manuals should be in English. Provision of trained manpower support for at least for initial 3 months after installation or up to a period when the staff/technicians at AIIMS Raipur are trained in handling, storage and proper sterilization methods of supplied equipment has to be ensured by the bidder.	User manuals should be in English. Provision of trained manpower support for at least for initial 1 months after installation or up to a period when the staff/technicians at AIIMS Raipur are trained in handling, storage and proper sterilization methods of supplied equipment has to be ensured by the bidder.
Page No. 18 & 22	Should have standard Accessories Monopolar Handle, Cables, Foot Paddle, Bipolar Forceps , Cable, Neutral Plate, Cable and Electrodes, irrigation unit, (if not	Should have standard Accessories Monopolar Handle, Cables, Single switch Foot Paddle for Bipolar and ,double switch foot paddle for monopolar,



	integrated) should have standard Accessories like power cord, pressure cut off & tube set.	Bipolar Forceps , Cable, Neutral Plate, Cable and Electrodes, irrigation unit, (if not integrated) should have standard Accessories like power cord, pressure cut off & tube set.
Page No. 18 & 22	Loop Monopolar electrodes tips of different sizes- 5 each (size 6mm, 12mm, and 16mm dia.)	Loop Monopolar electrodes tips of different sizes- 5 each (size 6mm, 12mm, 16mm dia.) with reusable pencil to use along with it.
Additional Point On specification of Annexure I & II	NIL	Added: Annexure-I: 1. Microprocessor Controlled Electro surgery unit 2. LCD display with focused view of current active mode. 3. Visual and audible alarms during activation 4. Storage of 10 user defined programmable profiles. 5. Facility to store up to 10 recently occurred error message 6. Should have continuous patient monitoring for neutral electrode. 7. Should have bipolar coagulation facility with Auto start facility. 8. Should support coagulation from two monopolar simultaneously 9. Should have universal socket for monopolar, bipolar and neutral electrodes. 10. One bipolar and two monopolar outputs.
	NIL	Added: Annexure-II: 1. Microprocessor Controlled Electro surgery unit 2. LCD display with focused view of current active mode. 3. Visual and audible alarms during activation 4. Storage of 10 user defined programmable profiles. 5. Facility to store up to 10 recently occurred error message 6. Should have continuous patient monitoring for neutral electrode. 7. Should have bipolar coagulation facility with Auto start facility. 8. Should support coagulation from two monopolar simultaneously 9. Should have universal socket for monopolar, bipolar and neutral



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

		electrodes. 10. One bipolar and two monopolar outputs.
On specification of Annexure I & II	NIL	1. Should have supplied instruments to AIIMS, PGI & NIMHANS 2. Should produce satisfactory performance certificate from respective institute.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)