



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़  
(All India Institute of Medical Sciences, Raipur (Chhattisgarh)  
Tatibandh, GE Road, Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

No. AIIMS/R/CS/045/EEG-ERP/PHY/18/OT

Date - 02/11/2018

### **Corrigendum**

**Sub: Corrigendum in Notice Inviting Tender for “64 Channel EEG-ERP System” for Department of Physiology, AIIMS Raipur.**

**Tender Id :- 2018\_IMSRP\_391763\_1 date 05/10/2018**

In context to following corrigendum is issued against subject:

S.no.	Page/Clause/Point	Existing	To be read as/Addenda
1.	Page 17 Clause - Detailed Specification	Input Noise < 0.5 $\mu$ V	Input Noise 0.5 $\mu$ V to 2 $\mu$ V
2	Page 17 Clause - Detailed Specification	Sampling Rate: up to 20000 Hz per channel using Wi-fi/ Optic fiber connectivity	Sampling Rate: to be between 8000 Hz to 20000 Hz per channel using Wi-fi/ Optic fiber connectivity.

**Note:**

- All other terms & condition will be remain unchanged.*
- Corrigendum must be uploaded along with technical bid as acceptance.*

(सुशील सोनबेर)  
भंडार अधिकारी  
अखिल भारतीय आयुर्विज्ञान संस्थान,  
रायपुर (छ.ग.)