

अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर (छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

Raipur, Dtd:- 05.07.2018

No. AIIMS/R/CS/OBGY/17/2565/Tender

<u>Corrigendum</u>

Tender ID.No.2018_IMSRP_344767_1, Dated:04.06.2018.

With reference to above Tender ID.No., the following amendment is issued:

Page/Clause/ Point	Existing Sentence	To be read as
Point.No.20	It should be provide trolley	Trolley is not required as cart is already mentioned in point number 18.
Point.No.15.	The system should be supplied with all the standard reusable accessories such as- Electrode set of 10 Nos. each (Pin point, flat tip-short, flat tip- long blade), Monopolar double pedal foot switch, earth pad – 5, bipolar forcep with cord -5, electrode handle with finger switch- 10 and other standard accessories It should be comfortable for under water surgery like hysteroscopy Open surgery reusable curved vessel sealing clamp-2 Laparoscopic reusable 5mm vessel sealing instrumenyt-1	 1.All the accessories which are being asked in the tender including monopolar pencils,Bipolar Forceps,Vessel Sealing Forceps,Bipolar Scissors,Cart System etc should be of the same manufacturer is to be added in the existing specifications. 2.The quantity of open surgery reusable curved vessel sealing clamp is hould be -4 Nos and quantity of laparoscopic reusable 5mm vessel sealing instrument is should be-4 Nos.
Point.No.16.	The unit should be having quality standard certified by DIN ISO 9001/ EN 46001.	The equipment comply mandatory with the requirement of the medical device directive of class I equipment and electronic agnetic compability certificates such as – IEC 6060 1-2 (Certificate for standard Electro Magnetic Compatibility) IEC 6060 1-1 (Certificate for Standard Safety requirement for medical electrical Equipment) is to be added in the existing specification.
New Points is added		Should have the option of Bipolar resection software upgradation and must be compatible with resect scopes from different manufacturers.

All other terms and condition will remain unchanged.

(सुशील सोनबेर) भंडार अधिकारी अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)