



No. AIIMS/R/CS/MS/Defibrillator/18/711/OT/A

Dtd:05/11/2019

Corrigendum

Subject:- Supply and Installation of Biphasic Defibrillator with ECG Machine for AIIMS Raipur.

Page	/Clause/Point No.	Existing	To be read as
Page No 16 of 20	Point no. 8	Should have external paddles with paddles contact indicator- for good paddle contact. Single adult and paediatric paddles should be available	Should have external paddles with paddles contact indicator- for good paddle contact. Single adult and paediatric paddles should be available with each unit
Page No 16 of 20	Point no. 33	Accessories: <ul style="list-style-type: none">• 5/6 lead ECG cable• AED pad for adult & Paediatric – 2 each	Accessories: <ul style="list-style-type: none">• 5/6 lead ECG cable• AED pad for adult & Paediatric – 2 each per defibrillator• Internal paddles with handle assembly -Adult -6, paediatric -6, infant-2• Infant external paddles-4
Page No 16 of 20	Clause no 28s	Selectable shock energy in AED mode	Selectable/fixed shock energy in AED mode
Page No 16 of 20	Point no. 30	Equipment must have MIL – STD-810F 514.5 category 4 restrained cargo certification exposure level annex A 2.2.1C(1) & MIL-STD-810F 514.5, category 9 helicopter EN78 9: 2007, Am1: 2010 certification	Equipment must have MIL – STD-810F 514.5 category 4 restrained cargo certification exposure level annex A 2.2.1C (1) and MIL-STD-810F514.5.

Note: - Vendor should quote the price of battery and AED pads as consumable for further replacement. Price for battery and AED pads will be freezed for next 5 yrs.

**Sr. Administrative Officer
AIIMS Raipur (CG)**