



क्रमांक: AIIMS/R/CS/Paedia/18/562/OT/Corrigendum/

दिनांक: 06/02/2019

Tender ID No.: 2019_IMSRP_428601_1

With reference to above tender ID No., the following amendment is being issued:-

Corrigendum 02

Page/Clause/ Point	Existing	To be read as
Page no. 19, point no 2	2. Technical: i. Distal end & insertion tube outer diameter should be 3 mm or less. iii. Field of view should be 90 degree vii. Insertion tube outer diameter 3.5 to 4.00 mm	i. Distal end & insertion tube outer diameter should be 4 mm or less. iii. Field of view should be at least 90 degree vii. Deleted
Page no. 19, point no 3	3 Video Processor for Bronchoscope v. Narrow Band Imaging capacity for compatibility with NBI Videoscopes	v. Narrow Band Imaging/iScan OE capacity for compatibility with Videoscopes
Page no. 19, point no 4	4. Light Source (Xenon short arc Ozone free 300 Watt lamp): i. Equipped with Narrow Band Imaging capability with high intensity Xenon Light source (300W) with 500 hours life. iii Equipped with special filter for Narrow Band Imaging. v. Emergency halogen light for backup.	i. Equipped with Narrow Band Imaging/iScan OE capability with high intensity Xenon Light source (300W) with 500 hours life. iii. Equipped with special filter for Narrow Band Imaging//iScan OE. v. Emergency halogen/LED light for backup.
Page no. 20, point no 7	7. Common Points ii. Comprehensive warranty for 5 years and 5 years CMC after warranty	ii. Comprehensive warranty for 5 years and 5 years CMC after warranty (Accidental damage should be covered in standard cost)

All other terms and condition will remain unchanged.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)