



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)

No. AIIMS/R/CS/PuI_Med/18/009/OT

Raipur, Dtd:-06/02/2019

Corrigendum

Group/ Point no of Technical Specification on Annexure - I	Existing Sentence	To be read as
Name of the Item	EBUS TBHA / BIOPSY SYSTEM	EBUS TBNA / BIOPSY SYSTEM
Group 1.1	Parameter (gain, baseline, PRF etc.) Dynamic Tissue harmonic imaging.	Parameter (gain, baseline, PRF etc.) latest Tissue harmonic imaging and tissue elastography.
Group 1.3.II	100-300 W xenon lamp (Lamp life 300-500 hrs) / LED light source with back up lamp.	Atleast 300W xenon lamp (lamp life 300 – 500 hrs) with backup lamp.
Group1.6. a	Warranty for five year and CMC for 5 year after the expiry of the warranty (CMC rates to be quoted separately)	Warranty for five year and CMC for 5 year after the expiry of the warranty including accidental cover for entire warranty period (CMC rates to be quoted separately)

Note :-

1. Technical specification remains the same as earlier except for above corrigendum.
2. All other terms and condition will remain unchanged.

**Stores Officer
AIIMS Raipur (CG)**