



AIIMS/R/CS/Neuro/80/18/OT/B/corrigendum

दिनांक:- 07.08.2019

Corrigendum

Tender ID No.: 2019_IMSRP_445932_2

With Reference to above tender ID No., the following amendment is being issued for the tender “**Ultra Sonic Surgical Aspirator**” for Neurosurgery Department at AIIMS Raipur:-

Annexure I

Existing		To be read as
S. no.	Specification Sr. No.	
1	2	The System should have following functional Subsystems: Fragmentation, Irrigation, Aspiration, Tissue Select Function and Electrosurgical Coagulation Capability
2	4	Main Unit:
		II. Based on the latest technology of Magneto Restrictive Technology.
3	5	Micro Surgical Hand piece with the system frequency between 30 to 40 KHz and weight of hand piece should not more than 80gm should be provided – 2 nos
4	6	Two operating frequency should be available which are suitable for removal of variety of tissue types like Soft Tissue, Calcified Tissue and Bone
5	7	Maximum tip amplitude should be 210 microns for micro hand piece.
6	14	Internal hand piece cooling system present for better functioning and longer life span, with cooling water flow of 35-50 cc/min.



7	15	The Suction path should be external so no contamination of hand piece takes place.	The Suction path should be external OR straight so no contamination of hand piece takes place.
8	25	<ul style="list-style-type: none">• Shear Tip with Titanium Nitride coating for fragmenting and cutting through fibrous and collagen-rich tissue such as fibrous meningiomas and connective tissue for micro hand piece – 1 no.	Shear Tip OR notched tip for fragmenting and cutting through fibrous and collagen-rich tissue such as fibrous meningiomas and connective tissue for micro hand piece – 1 no.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)