

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

दिनांक: 09/10/2018

क्रमांकः AIIMS/R/CS/Pedia S./18/243/OT/Corrigendum/

Tender ID No.: 2018_IMSRP_376689_1

With reference to above tender ID No., the following amendment is being issued:-

Corrigendum -02

Page/Clause/	Existing	To be read as
Point		
Page no. 18, point no. 3	It should have HL7 gateway to connect the HIS, CIS.	It should have HL7 connectivity to connect the HIS, CIS, central monitor. (Connectivity to central monitor to be done by vendor)
Page no. 18, point no 4	Monitor should have facility for Monitoring the following parameters – ECG,Respiration,SpO2/NIBP,2Temp,2IBP,Main stream ETCO2 ,Drug calculation,Lung function Hemodynamic data and Oxy CRG	Monitor should have facility for Monitoring the following parameters –ECG, Respiration ,SpO2,NIBP, 2 Temp, 2 IBP ,Main stream /Side stream ETCO2 , Drug calculation ,Lung function ,Hemodynamic data and Oxy CRG.
Page no. 18, point no 11	Monitor should have facility to store minimum 1600 arrhythmia episode.	Monitor should have facility to store minimum 400 Snapshots and 800 event storage.
Page no. 18, point no 12	Monitor should have facility for full Disclosure of ECG and 4 other parameter of last 24 hours.	Monitor should have facility for Event storage and snapshot facility for last 24 hours.
Page no. 18, point no 19	SpO2: Measurements technology: Masimo/Nihon Kohen/Fast. Measurement Method: Two wavelength light absorption Method. Measurement range: 0-100% Accuracy Adults: 70-100% +/- 2 digit, 0-69% unspecified. Accuracy Neonates: 70-100% +/- 3 digits, 0-69% unspecified, Pulse rate range: 20 to 300 bpm	SpO2: Measurements technology: Masimo SET Measurement Method: Two wavelength light absorption Method. Measurement range: 0-100% Accuracy Adults: 70-100% +/- 2 digit, 0-69% unspecified. Accuracy Neonates: 70-100% +/- 3 digits, 0-69% unspecified, Pulse rate range: 20 to 300 bpm
Page no. 19, point no 30	Alarm trend (Recall): Minimum 8000 events	Alarm Trend (Recall): minimum 400 Snapshots and 800 events.
Page no. 19, point no 35	Reusable Pediatric, Neonatal probe -2each	Reusable Pediatric, Neonatal SpO2 probe - 2each
Page no. 19, point no 36	ECG cable reusable – 2sets	ECG cable reusable 3/5 lead with ESU, Defibrillator protection - 1 set, 10 Lead ECG - 1set
Additional points		Reusable temperature probe -5 to be supplied

All other terms and condition will remain unchanged.

(सुशील सानबेर) भंडार अधिकारी अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)