



अखिल भारतीय आयर्वजिन सस्थान ,रायपुर (छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

No. AIIMS/R/CS/21149/Oph/19/OT/A

Date – 09/12/2019

Corrigendum

Sub: Corrigendum in Notice Inviting Tender for “Supply of Retina Laser(Multi Spot) with SL,IO and Endo Delivery System” for Department of Ophthalmology, AIIMS Raipur.

Tender Id: - 2019_IMSRP_521341_1 date 23/11/2019 In context to following corrigendum is issued against subject:

S.no.	Page/Clause/Point	To be added
1.	Page 19 (on Technical Specifications to	<ul style="list-style-type: none">• PAR FOCUS TECHNOLOGY• Insight view display: Micro display integrated in LSL. Surgeon need not to check his parameters outside or by moving his eye during the laser procedure. Surgeon can stay focused & can see all the parameters on the right eye piece.• Fixed clear view PSF (physician safety Filter).• Automated laser treatment summary report: Detailed reports are generated after successfully completion of patient’s treatment. As a result surgeon receive a complete glance of Laser Treatment summary report in touch control panel.• Smaller in volume lighter in weight fibre ports.• Upgradation to selective Laser Trabeculoplasty (SLT) for Glaucoma.
2.	Page 19/Point no.11/ (on Technical Specifications typographical error)	. Laser safety eye glasses and endo laser filters for standard operating microscope.

Note:

- All other terms & condition will be remain unchanged.*
- Corrigendum must be uploaded along with technical bid as acceptance.*

**Sr.Administrative Officer
AIIMS Raipur**